अखिल भारतीय आयुर्विज्ञान संस्थान, रायपुर (छत्तीसगढ़)

All India Institute of Medical Sciences, Raipur (Chhattisgarh)



Tatibandh, GE Road, Raipur 492 099 (CG)

www.aiimsraipur.edu.in

No.: AIIMS/R/CS/PC/19/29/LPC/A

QUOTATION NOTICE

Date: 23-05-2020

Inviting quotations for purchase of "20 watt LED batten fittings" for Department of Project cell at AIIMS Raipur.

Sealed quotations are invited from intending registered Stockiest / Distributors having GST and relevant documents for above said. The quotation should be submitted to, Medical College Building, 2nd Floor, gate no.-5, AIIMS, Raipur office of the Stores Officer up to **01-06-2020** before **3:00** pm. The quotations will be opened on the same day at **3.30 pm**.

Terms & Condition:

- 1. Quotation should be in **two- bid system** i.e. **Technical Bid** (as per **Annexure-I**) and **Financial Bid** (as per **Annexure-II**).
- 2. The technical bid and the financial bid should be sealed by the bidder in separate covers which will be super-scribed "Technical Bid for 20 watt LED batten fittings" and "Financial Bid for 20 watt LED batten fittings ". Both Sealed Envelopes should be kept in a main/ bigger envelope super-scribed as "Quotation for supply of 20 watt LED batten fittings".
- **3.** All the required documents should be enclosed in the envelope of technical bid and financial bid should be enclosed in the envelope of financial bid.
- 4. Rate should be mentioned in words & figure both.
- 5. Taxes/GST, if any (Kindly mention in table)
- 6. Delivery schedule: within 15 days from the date of issue of P.O.
- 7. No additional documents related to this NIQ will be entertained after opening of NIQ.
- 8. Price should be FOR Destination basis (i.e. Department of project cell, AIIMS Raipur).
- 9. LD @ 0.5% of delayed supply per week or part thereof for delay of supply of material subject to maximum up to 10% should be deducted. After expiry of delivery period material cannot be accepted without extension of delivery period.
- 10. Quotation No/Name and due date of opening must be written on top of envelop.
- 11. Brand, Make & Warranty should be clearly mentioned in offer as well as tender/ quotation. Specific authorization may be submitted with the offer/bid.
- 12. GST rates applicable on your quoted item may please be confirmed. **HSN code for each item should be clearly mentioned**.
- 13. Any change (Upward/Reduction) in the Basic Price structure post submission of bid shall be promptly intimated to AIIMS Raipur. You are also required to pass the Input Credit as per the following **Anti Profiteering Clause of GST**.

"Upon Implementation of GST, any reduction in the rate of tax on supply of goods or service or the benefit of input tax credit shall be passed on to AIIMS Raipur by way of commensurate reduction in the prices".

- 14. The GST registration details may be furnished.
- 15. RTGS details required for payment purpose.
- 16. 100% payment against receipt and acceptance of material.
- 17. No part supply or part payment will be entertained.
- 18. Validity of offer should not be less than 90 days.
- 19. Supply, installation and commissioning will be done by firm (if applicable).
- 20. The quantity in above column is totally tentative. It can be increased or decreased at the time of placement of order.
- 21. AIIMS, Raipur reserve the right to place the order for full or part quantity to one or more firms.



अखिल भारतीय आयुर्विज्ञान संस्थान, रायपुर (छत्तीसगढ़) All India Institute of Medical Sciences, Raipur (Chhattisgarh) Tatibandh, GE Road, Raipur-492 099 (CG) www.aiimsraipur.edu.in

ANNEXURE-I

Ref. No. & Date: -Quotation Notice No.: -Brand/Make Name:-Model No.:-Warranty-

TECHNICAL BID

Purchase of "20 watt LED batten fittings" for Department of Project cell at AIIMS Raipur. Details of item are given as under:

S. N.	Description of item(s)	Qty./ Unit	Offered Specification	Specify Technically Complied / Not Complied/ if any Deviations
<u>N.</u> 1.	 20 watt LED batten fittings: Colour temperature 6000 K± 500 K Ingress protection IP20 warranty for free replacement 2 year LED lamp efficiency (Lm/W) 100 ±10% Length of tube light (nominal) (in mm): - 1220 Colour Rendering Inde (CRI):>/= 80 Beam Angle (degree):-> /= 120 degree operating voltage range (volt) :-100-300 Drive efficiency (%) :- more than 85% Surge voltage withstand (KV):- 2.5 KV Power factor :-> /0.9 Type of tube light cap:- G-13 Bi-pin lamp shell material :- Polycarbonate. Approved make - Wipro, Bajaj ,CGL, Phillips , Osram Equivalent 	500 Nos.		Complied/ if any Deviations
	Warranty-02 year			

Sign of bidder:-Date:-Name of the bidder:-Seal of the bidder:-:-



ANNEXURE-II

Ref. No. & Date: -Quotation Notice No.: -

FINANCIAL BID

Purchase of "20 watt LED batten fittings:" for Department of Project cell at AIIMS Raipur. Details of item are given as under:

-	item are given as under.							
S.	Description of item(s)	Qty./	HSN	Make/	Unit	GST Per	Total Unit	Total
N.		Unit	Code	Brand/	Price in	Unit @	Price with	Amount in
				Model	₹.		Tax in ₹.	₹.
1.	20 watt LED batten fittings:							
	Color temperature $6000 \text{ K} \pm 500 \text{ K}$	500						
	Ingress protection IP20	Nos.						
	warranty for free replacement 2							
	year							
	LED lamp efficiency (Lm/W) 100							
	±10%							
	Length of tube light (nominal) (in							
	mm): - 1220							
	Color Rendering Inde (CRI):>/=							
	80							
	Beam Angle (degree):-> /= 120							
	degree							
	operating voltage range (volt) :-							
	100-300 Drive efficiency (%) :-							
	more than 85% Surge voltage							
	withstand (KV):- 2.5 KV							
	Power factor :-> $/0.9$							
	Type of tube light cap :- G-13 Bi-							
	pin lamp shell material :-							
	Polycarbonate.							
	Approved make-Wipro, Bajaj							
	,CGL, Phillips ,Osram Equivalent							
	Authorization certificate from							
	OEM							
	Warranty-02 year							

Sign of bidder:-Date:-Name of the bidder:-Seal of the bidder:-