



No.: AIIMS/R/CS/Patho/20/24/LPC/

Date: 08.06.2020

QUOTATION NOTICE

Inviting quotations for purchase of “**Wax Trimmer**” for Department of Pathology & Lab Medicine at AIIMS Raipur.

Sealed quotations are invited from intending registered Stockiest / Distributors having GST and relevant documents for above said. The quotation should be submitted to, Medical College Building, 2nd Floor, gate no.-5, AIIMS, Raipur office of the Stores Officer up to **15.06.2020** before **3:00 pm**. The quotations will be opened on the same day at **3.30 pm**.

Terms & Condition:

1. Quotation should be in **two- bid system** i.e. **Technical Bid** (as per **Annexure-I**) and **Financial Bid** (as per **Annexure-II**).
2. The technical bid and the financial bid should be sealed by the bidder in separate covers which will be super-scribed “**Technical Bid for Wax Trimmer**” and “**Financial Bid for Wax Trimmer**”. Both Sealed Envelopes should be kept in a main/ bigger envelope super-scribed as “**Quotation for supply of Wax Trimmer**”.
3. **All the required documents should be enclosed in the envelope of technical bid and financial bid should be enclosed in the envelope of financial bid.**
4. Rate should be mentioned in words & figure both.
5. Taxes/GST, if any (Kindly mention in above table)
6. Delivery schedule: within 15 days from the date of issue of P.O.
7. No additional documents related to this NIQ will be entertained after opening of NIQ.
8. Price should be FOR Destination basis (i.e. Department of Pathology, AIIMS Raipur).
9. LD @ 0.5% of delayed supply per week or part thereof for delay of supply of material subject to maximum up to 10% should be deducted. After expiry of delivery period material cannot be accepted without extension of delivery period.
10. Quotation No/Name and due date of opening must be written on top of envelop.
11. **Brand, Make & Warranty should be clearly mentioned in offer as well as tender/ quotation. Specific authorization may be submitted with the offer/bid.**
12. GST rates applicable on your quoted item may please be confirmed. **HSN code for each item should be clearly mentioned.**
13. Any change (Upward/Reduction) in the Basic Price structure post submission of bid shall be promptly intimated to AIIMS Raipur. You are also required to pass the Input Credit as per the following **Anti Profiteering Clause of GST**.
“Upon Implementation of GST, any reduction in the rate of tax on supply of goods or service or the benefit of input tax credit shall be passed on to AIIMS Raipur by way of commensurate reduction in the prices”.
14. The GST registration details may be furnished.
15. RTGS details required for payment purpose.
16. 100% payment against receipt and acceptance of material.
17. No part supply or part payment will be entertained.
18. Validity of offer should not be less than 90 days.
19. Supply, installation and commissioning will be done by firm (if applicable).
20. The quantity in above column is totally tentative. It can be increased or decreased at the time of placement of order.
21. AIIMS, Raipur reserve the right to place the order for full or part quantity to one or more firms.
22. Due of CVOID-19 soft copy of quotation notice is acceptable on the company letterhead with seal and signature in storesofficer.cp@aiimsraipur.edu.in mail id up to 15.06.2020 before 3:00 pm.
23. The product should be of superior quality and highly standard.

Stores Officer,
AIIMS, Raipur (C.G.)



ANNEXURE-I

Ref. No. & Date: -

Quotation Notice No.: -

Brand/Make Name:-

Model No.:-

Warranty-

TECHNICAL BID

Purchase of “**Wax Trimmer**” for Department of Pathology & Lab Medicine at AIIMS Raipur. Details of item are given as under:

S. N.	Description of item(s)	Qty./Unit	Offered Specification	Specify Technically Complied/Not Complied / if any Deviations
1.	The trimmer should be designed to remove excess of paraffin from the embedding cassettes after inclusion of specimens.			
2.	Scraping the embedded block on the "grooved" plane heated surface should remove excess paraffin leaving the cassette clean and ready to be set properly on the clamp of the microtome.			
3.	The excess of melted paraffin should be collected in a practical removable drawer.			
4.	Temperature range of operation should be from 30° to 90° C			
5.	Precision in the control of temperature should be $\pm 1^\circ\text{C}$			
6.	Power supply up to 220V/50Hz			
7.	Dimension & weight should be (around): Width: 22cm Depth: 16cm Height: 17.5cm and Weight: 2.5 Kg.			
8.	Document supporting track record and satisfactory performance from institutes of national importance (minimum one) should be provided.			
9.	Support for induction and follow up training of technical staff, on-site standardization and troubleshooting of procedures/tests to be provided by the company.			
10.	The system should comply with European CE, US FDA or BIS approved.			
11.	02 years warrantee should be provided.			

Sign of bidder :-

Date :-

Name of the bidder :-

Seal of the bidder :-



ANNEXURE-II

Ref. No. & Date: -

Quotation Notice No.: -

FINANCIAL BID

Purchase of “**Wax Trimmer**” for Department of Pathology & Lab Medicine at AIIMS Raipur. Details of item are given as under:

S. N.	Description of item(s)	Qty./ Unit	Offered Specification	Specify Technically Complied/Not Complied / if any Deviations
1.	The trimmer should be designed to remove excess of paraffin from the embedding cassettes after inclusion of specimens.			
2.	Scraping the embedded block on the "grooved" plane heated surface should remove excess paraffin leaving the cassette clean and ready to be set properly on the clamp of the microtome.			
3.	The excess of melted paraffin should be collected in a practical removable drawer.			
4.	Temperature range of operation should be from 30° to 90° C			
5.	Precision in the control of temperature should be $\pm 1^\circ\text{C}$			
6.	Power supply up to 220V/50Hz			
7.	Dimension & weight should be (around): Width: 22cm Depth: 16cm Height: 17.5cm and Weight: 2.5 Kg.			
8.	Document supporting track record and satisfactory performance from institutes of national importance (minimum one) should be provided.			
9.	Support for induction and follow up training of technical staff, on-site standardization and troubleshooting of procedures/tests to be provided by the company.			
10.	The system should comply with European CE, US FDA or BIS approved.			
11.	02 years warrantee should be provided.			

Sign of bidder :-

Date :-

Name of the bidder :-

Seal of the bidder :-