



No.: AIIMS/R/CS/MOT/Capital/19-20/59/

Date:18/07/2020

QUOTATION NOTICE

Inviting quotations for purchase of “Patient Convective Warming Device” for MOT B Block 4th Floor at AIIMS Raipur.

Sealed quotations are invited from intending registered Stockiest / Distributors having GST and relevant documents for above said. The quotation should be submitted to, Medical College Building, 2nd Floor, gate no.-5, AIIMS, Raipur office of the Stores Officer or submitted through on Email – storesofficer.cp@aiimsraipur.edu.in up to 24/07/2020 before 3:00 pm. The quotations will be opened on the same day at 3.30 pm.

Terms & Condition:

1. Quotation should be in **two- bid system** i.e. **Technical Bid** (as per **Annexure-I**) and **Financial Bid** (as per **Annexure-II**).
2. The technical bid and the financial bid should be sealed by the bidder in separate covers which will be super-scribed "**Technical Bid for Patient Convective Warming Device**" and "**Financial Bid for Patient Convective Warming Device**". Both Sealed Envelopes should be kept in a main/ bigger envelope super-scribed as "**Quotation for supply of Patient Convective Warming Device**".
3. **All the required documents should be enclosed in the envelope of technical bid and financial bid should be enclosed in the envelope of financial bid.**
4. Rate should be mentioned in words & figure both.
5. Taxes/GST, if any (Kindly mention in table)
6. Delivery schedule: within 15 days from the date of issue of P.O.
7. No additional documents related to this NIQ will be entertained after opening of NIQ.
8. Price should be FOR Destination basis (i.e. MOT B Block 4th Floor, AIIMS Raipur).
9. LD @ 0.5% of delayed supply per week or part thereof for delay of supply of material subject to maximum up to 10% should be deducted. After expiry of delivery period material cannot be accepted without extension of delivery period.
10. Quotation No/Name and due date of opening must be written on top of envelop.
11. **Brand, Make & Warranty should be clearly mentioned in offer as well as tender/ quotation. Specific authorization may be submitted with the offer/bid.**
12. GST rates applicable on your quoted item may please be confirmed. **HSN code for each item should be clearly mentioned.**
13. Any change (Upward/Reduction) in the Basic Price structure post submission of bid shall be promptly intimated to AIIMS Raipur. You are also required to pass the Input Credit as per the following **Anti Profiteering Clause of GST.**
“Upon Implementation of GST, any reduction in the rate of tax on supply of goods or service or the benefit of input tax credit shall be passed on to AIIMS Raipur by way of commensurate reduction in the prices”.
14. The GST registration details may be furnished.
15. RTGS details required for payment purpose.
16. 100% payment against receipt and acceptance of material.
17. No part supply or part payment will be entertained.
18. Validity of offer should not be less than 90 days.
19. Supply, installation and commissioning will be done by firm (if applicable).
20. The quantity in above column is totally tentative. It can be increased or decreased at the time of placement of order.
21. AIIMS, Raipur reserve the right to place the order for full or part quantity to one or more firms.

**Stores Officer,
AIIMS, Raipur (C.G.)**



ANNEXURE-I

Ref. No. & Date: -
Quotation Notice No.: -
Brand/Make Name:-
Model No.:-
Warranty-

TECHNICAL BID

Purchase of “**Patient Convective Warming Device**” for MOT B Block 4th Floor at AIIMS Raipur. Details of item are given as under:

S. N.	Description of item(s)	Qty./ Unit	Offered Specification	Specify Technically Complied / Not Complied/ if any Deviations
1.	Patient Convective warming Device : 1.Dimension: 29-30 cm x 22-24 cm x 38 - 40 cm. 2.Wright : 6-7 kg 3.Operating Temperatures: Ambient, 34 C , 37 C , 40 C , 46 C +/- 2 C with delivery hose cover. 4.Fan Speed and Relative Noise: 38 -40 CFM (AIR FLOW) / < 50 dBM (NOISE LEVEL) 5.Leakage Current: Meets UL 60601 – 1 3 rd edition and IEC 60601 – 1 3 rd Ed requirements. 6.Safety System: Dual temperature limit thermostats / heater temp. Limit devices / temperature limit devices / Control system failure alarm. 7.Filter: Bacteriological HEPA filter/ change after 1000 hour of use. 8.Device Rating 1050 – 100 Watts Maximum. 9.Compatibility : ALL type make blanket full body (05 Nos. of Blanket) 10.Power cord, All manual soft copy in Pen drive, Connecting cable for PC , and other machine accessories compatible with Indian sockets. 11.Warranty for 02 years. 12.System should be BIS/ISO/European CE/ USFDA Approved. Certificate of standard should be attached with tender.	01 Nos.		

Sign of bidder :-
Date :-
Name of the bidder :-
Seal of the bidder :-



ANNEXURE-II

Ref. No. & Date: -

Quotation Notice No.: -

FINANCIAL BID

Purchase of “**Patient Convective Warming Device**” for MOT B Block 4th Floor at AIIMS Raipur. Details of item are given as under:

S. N.	Description of item(s)	Qty./ Unit	HSN Code	Make/ Brand/ Model	Unit Price in ₹.	GST Per Unit @	Total Unit Price with Tax in ₹.	Total Amount in ₹.
1.	Patient Convective warming Device : 1.Dimension: 29-30 cm x 22-24 cm x 38 - 40 cm. 2.Wright : 6-7 kg 3.Operating Temperatures: Ambient, 34 C , 37 C , 40 C , 46 C +/- 2 C with delivery hose cover. 4.Fan Speed and Relative Noise: 38 -40 CFM (AIR FLOW) / < 50 dM (NOISE LEVEL) 5.Leakage Current: Meets UL 60601 – 1 3 rd edition and IEC 60601 – 1 3 rd Ed requirements. 6.Safety System: Dual temperature limit thermostats / heater temp. Limit devices / temperature limit devices / Control system failure alarm. 7.Filter: Bacteriological HEPA filter/ change after 1000 hour of use. 8.Device Rating 1050 – 100 Watts Maximum. 9.Compatibility : ALL type make blanket full body (05 Nos. of Blanket) 10.Power cord, All manual soft copy in Pen drive, Connecting cable for PC , and other machine accessories compatible with Indian sockets. 11.Warranty for 02 years. 12.System should be BIS/ISO/European CE/ USFDA Approved. Certificate of standard should be attached with tender.	01 Nos.						

Sign of bidder :-

Date :-

Name of the bidder :-

Seal of the bidder :-