



No.: AIIMS/R/CS/Anaesthesiology/0444/2022/

Date: 13-09-2022

Quotation notice for the purchase of Ozone generator Analog Dual Feed-Esensor for Anaesthesiology Department at AIIMS, Raipur under GFR 155

Inviting quotations purchase of Ozone generator Analog Dual Feed-Esensor for Anaesthesiology Department at AIIMS, Raipur.

Sealed quotations are invited from intending registered Stockiest/ Distributors having GST and relevant documents for above said product. The quotation should be submitted to Medical College Building, ground floor, gate no.-5, AIIMS, Raipur office of the Stores Officer up to **20/09/2022** before **3:00 pm**. The quotations will be opened on the same day at **3.30 pm**. Details of item are given as under:-

S. N.	Description of item(s)	Make	Qty./ Unit (Nos.)	unit price	GST %	unit price with GST	Total Price
1.	<p>Ozone generator Analog Dual Feed-Esensor:</p> <ul style="list-style-type: none">- Should be a digital model- Should have real time on line calibration visible on screen- Should be able to be used in pain therapy in spine and arthritis<ul style="list-style-type: none">a. Ozone Discolysisb. Ozone nucleotomyc. Arthritic injection- Should be microprocessor controlled- Must have self cooled design- Must be light weight- Should be easily adjustable at each 1 mcg division from 7 to 98 mcg/ ml- Should have perfect calibration of output- Should have internal circuit technology has to be climate-compatible.- Machine has to remain cool as long as Ozone is produced- Must have auto adjustments according to variable ambient conditions.- Should be having stability of manufactured Ozone for atleast 1000 Seconds.- Absolute ozone output should be at least 2000 mg per hour- Output calibration should be up to atleast 70 micrograms		1 no				



अखिल भारतीय आयुर्विज्ञान संस्थान रायपुर (छत्तीसगढ़)
All India Institute of Medical Sciences, Raipur (Chhattisgarh)

Tatibandh, GE Road, Raipur-492 099 (CG)

www.aiimsraipur.edu.in

	per ml. - Output should be capable of adjusting from outside with 1 microgram per division. - Warranty-1 year						
--	---------------------------------------------------------------------------------------------------------------------	--	--	--	--	--	--

Terms & Condition:

- 1) Rate should be mentioned in words & figure both, Product brochure should be attached with quotation.
- 2) Delivery Schedule – within 15 days from the date of issue of PO.
- 3) Price should be FOR Destination basis. (i.e Anaesthesiology Department, AIIMS Raipur).
- 4) LD @0.5% of delayed supply per week or part of week for delay of supply of material subject to maximum up to 10%.
- 5) After expiry of delivery Period material cannot be accepted without the extension of delivery period.
- 6) Quotation No/Name and Due date of opening must be written on top of envelope.
- 7) GST rates applicable on your quoted item may please be confirmed. HSN code for each item shown be clearly mentioned.
- 8) Please confirm if there any change (Upward/Reduction) in your Basic Price structure. And you are also requested to pass the Input Credit as per the following Anti Profiteering Clause of GST **“Upon Implementation of GST, any reduction in the rate of tax on supply of goods or service or the benefit of input tax credit shall be passed on to AIIMS Raipur by way of commensurate reduction in the prices”**.
- 9) The GST registration details may please be furnished.
- 10) 100% payment against receipt and acceptance of material.
- 11) Validity of offer should not be less than 90 days.
- 12) RTGS details required for payment purpose.
- 13) No part supply or part payment will be entertained.
- 14) Expenditure will be debitable to GIA-GENERAL.
- 15) Brand, Make & warranty should be clearly mentioned in offer (if required)
- 16) The Quantity of above column is totally tentative. It can be increased or decreased at the time of placement of order.
- 17) Supply, installation & commissioning will be done by Firm.
- 18) The product should be of superior quality and highly standard.

**Sr Proc Cum Stores Officer,
AIIMS, Raipur (C.G.)**