



क्रमांक/No.AIIMS/R/CS/URO/2023/28/LPC

दिनांक: 01/03/2023

Notice for inviting Quotation for Uroflowmeter System with Auto-clean facility for Urology Department at AIIMS Raipur.

कोटेशन सूचना
QUOTATION NOTICE

Sealed quotations are invited from intending registered Stockiest/ Distributors having GST/ relevant documents for supply Uroflowmeter System with Auto-clean facility for Urology Department at AIIMS, Raipur should be submitted to Medical College Building, Ground Floor, Gate no. 05 office of Store Officer up to 3:00 pm on **03 March 2023**, item description as per detailed bellow.

S. N.	Description of Item	Qty.	HSN Code	Make/ Brand/ Model	Unit Price in ₹	GST in %	Unit Price Incl. GST	Total Price in ₹.
1.	Uroflowmeter System with Auto-clean facility	1 Set						

नियम एवं शर्तें:

Terms & Condition:

1. Rate should be mentioned in word and figures both.
2. Taxes/GST, if any (Should be clearly mentioned).
3. Delivery Schedule – within 45 days from the date of issue of PO.
4. No any additional documents related to this NIQ will be entertained after opening of NIQ.
5. Price should be F.O.R. Destination basis (Urology Department).
6. LD @ 0.5% of delayed supply per week or part of week for delay of supply of material subject to maximum up to 10% of delayed supply should be deducted.
7. Quotation No/Name and Due date of opening must be written on top of the envelop.
8. 100% payment against received and acceptance of material.
9. No claim will be entertained regarding interest on any payment.
10. Where there is a statutory requirement for tax deduction at source, such deduction towards income tax and other tax as applicable will be made from the bills payable to the Supplier at rates as notified from time to time.
11. No payment shall be made for rejected stores. Rejected equipment's must be removed by the supplier within two weeks of the date of issue of rejection advice at their own cost & replace immediately. In case these are not removed these will be auctioned at the risk and responsibility of the suppliers without notice.
12. **Brand and make should be clearly mentioned in offer as well as tender/quotation specific authorization may be submit with the offer/Bid.**

13. GST rate applicable on your quoted item may please be confirmed. HSN code for each item should be clearly mentioned.
14. Please confirm if there any change (Upward/Reduction) in your Basic Price structure. And you are also requested to pass the Input Credit as per the following Anti Profiteering Clause of GST.
“Upon Implementation of GST, any reduction in the rate of tax on supply of goods or service or the benefit of input tax credit shall be passed on to AIIMS Raipur by way of commensurate reduction in the prices”.
15. In the event of increase in price, detailed justification and supporting evidence may be submitted.
16. The GST registration details may please be furnished.
17. AIIMS Raipur reserves the right to place the order for full or part quantity to one or more items.
18. RTGS detail required for payment purpose.
19. No part supply of part payment will be entertained.
20. Validity of offer should not be less than 90 days.
21. It should have warranty of **3 years**.
22. Supply, installation and commissioning will be done by firm (if applicable).
23. The quantity shown in above requirement column are totally tentative, it can be increase and decrease at the time of placement of purchase order.
24. AIIMS Raipur reserve the right to ask the sample for evaluation before placement of order.

Sr. Proc. Cum Store Officer
AIIMS, Raipur (C.G.)

Annexure-I

Sl. No.	Item Description	Qty.
1.	<p><u>Specifications for Uroflowmeter System with Auto-clean Facility:</u></p> <ol style="list-style-type: none">1. Flow system should have weight based gravimetric type uroflow transducer.2. There system should have auto-clean facility after patient urinate into the urine receiver with transducer. There should not be any need to manually clean after each test.3. Should have a flow range of 0-50 ml/sec with volume range of 0-1000 ml.4. Uroflowmeter system should be supplied with the micturition chair for voiding.5. The Uroflowmeter sensor unit should be connected via wire to the control unit placed in another room approximately 30 meters apart.6. The Control Unit should include database software, uroflow software, viewing monitor, keyboard, mouse and PC based printer for printing report.7. Should have the facility of automatic start, automatic stop of investigation and analysis by control Unit.8. Should have auto-record and zero facility.9. Sampling rate, reporting parameters and accuracy of the readings must be as per the ICS (International Continence Society) guidelines.10. Report format must contain all the report parameters required for the uroflowmetry diagnosis like: Patient identifications, position of voiding, hesitancy, Voided Volume, Qmax, Avg. flow, total flow time.11. Report format must contain required graphs for the report of uroflowmetry diagnosis like:<ol style="list-style-type: none">(a) Time vs Vmic (ml)(b) Time vs Qura (ml/sec)(c) ICS Nomogram curve for Qmax(d) ICS Nomogram curve for Average flow.12. The system must include all connection and power accessories required for full functioning of Uroflowmeter unit.13. The Uroflowmeter system should be provided with 03 (three) years warranty.	01 Set

**Sr. Proc. Cum Store Officer
AIIMS Raipur (C.G.)**