



No. AIIMS/R/HS/Surg_Gastro/2023/PAC

Date: 24/03/2023

Inviting quotations for procurement of Open Clip Applicator & Clips set for the department of Surgical Gastroenterology at AIIMS Raipur under Proprietary Article.

NOTICE INVITING QUOTATION

Sealed quotations are invited from intending registered Stockist / Distributors/traders having GST and relevant documents for procurement of Open Clip Applicator & Clips set for the department of Surgical Gastroenterology at AIIMS Raipur under Proprietary Article. The quotation with copy of certificate of GST & other documents should be submitted **through CPPP only upto 30/03/2023 before 03:00 pm.** The quotations/Bids will be opened on the next day i.e. on **31/03/2023 at 03:30pm.** Details of item are given as under:-

Sno.	Specification	Quantity Required	Unit	HSN Code	Brand	Unit Rate	GST			Amount
							SGST	IGST	CGST	
1.	Open Clip Applicator & Clips Set	01	Set							
The Bidders must submit the Purchase order copies of the same item supplied to other Central Government institutes.										
Grand Total										

Terms & Conditions:-

1. Firm to mention Make/Brand name in their quotation.
2. Taxes, if any (Kindly mention in above table) should be clearly mentioned in the offer.
3. Document relating to registration of firm i.e. GST and relevant document should be submitted along with quotation.
4. Products are certified from ISI/ISO/CE/GMP/BIS as applicable, the Certificate to this effect should be attached.
5. Supply should be done within 21 days after Placement of PO.
6. Price should be FOR Destination basis (i.e. concerned department).
7. 100% Payment will be released after certification from concerned department.
8. LD @ 0.5% of delayed supply per week or part of week for delay of supply of material subject to maximum up to 10% of delayed supply to be deducted.
9. AIIMS Raipur reserves the right to place order for full or part quantity to one or more firms. The AIIMS, Raipur reserves the right to increase/decrease the number of required quantity.

10. All other terms & condition as per GFR 2017.
11. Material to be delivered at **Surgical Gastroenterology Department, AIIMS Raipur.**
12. **Validity of the quotation should be 90 days from the date of opening.**
13. **Bidders should submit their email id, contact details with GST registration.**
14. **The Bidders must submit the Purchase order copies of the same item supplied to other institutes.**

Stores Officer (H)
AIIMS, Raipur (C.G)

Annexure –I

Open clip applicator & Clips set

Consists of Items:-

Sl. No.	Items	quantity required
1	Open Clip Applicator for clip size - small (blue) 7.5 inch	2
2	Open Clip Applicator for clip size - medium (silver) 7.5 inch	2
3	Open Clip Applicator for clip size - Medium/Large (green) 7.5 inch	1
4	Open Clip Applicator for clip size - Medium/Large (green) 10.5 inch	1
5	Open Clip Applicator for clip size - Large (yellow) 7.5 inch	1
6	Open Clip Applicator for clip size - Large (yellow) 10.5 inch	1
7	Titanium clips small - Pre-Sterilised Ligating clip cartridges, LT100 blue	36 cartridge (1 box)
8	Titanium clips medium Pre-Sterilised Ligating clip cartridges, LT200 silver	72 cartridge (2 box)
9	Titanium clips medium/large size Pre-Sterilised Ligating clip cartridges, LT300 green	36 cartridge (2 box)
10	Titanium clips Large size Pre-Sterilised Ligating clip cartridges, LT400 Yellow	18 cartridge (1 box)

Specification for Open Clip applicator and Clips set

1. Open Clip Applicator for clip size - small (blue) 7.5 inch:

- it should be:
 - i. Length 7.5 inch
 - ii. Could be used in Open procedures
 - iii. Applicable for small size (blue) clips (LT100)
 - iv. Tip angulated 15 degree with facility for distal tip closure and without ratchet
 - v. ETO Sterilizable
 - vi. Can be used in Multiple times (reusable)
 - vii. USFDA Certified

2. Titanium clips small - Pre-Sterilised Ligating clip cartridges, LT100 blue:

- One piece set must comprise of minimum 6 such clips.
- It should be compatible with the above mentioned clip applicator and both these items must be from the same manufacturer. (USFDA certified)

3. Open Clip Applicator for clip size - medium (silver) 7.5 inch:

- It should be:
 - i. Length 7.5 inch
 - ii. Could be used in Open procedures
 - iii. Applicable for medium size (silver) clips (LT200)
 - iv. Tip angulated 15 degree with facility for distal tip closure and without ratchet
 - v. ETO Sterilizable
 - vi. Can be used in Multiple times (reusable)
 - vii. USFDA Certified

4. Titanium clips medium Pre-Sterilised Ligating clip cartridges, LT200 silver:

- One piece set must comprise of minimum 6 such clips.
- It should be compatible with the above mentioned clip applicator and both these items must be from the same manufacturer. (USFDA certified)

5. Open Clip Applicator for clip size - Medium/Large (green) 7.5 inch:

- It should be:
 - i. Length 7.5 inch
 - ii. Could be used in Open procedures
 - iii. Applicable for Medium/Large size (green) clips (LT300)
 - iv. Tip angulated 15 degree with facility for distal tip closure and without ratchet
 - v. ETO Sterilizable
 - vi. Can be used in Multiple times (reusable)
 - vii. USFDA Certified

6. Open Clip Applicator for clip size - Medium/Large (green) 10.5 inch:

- It should be:
 - i. Length 10.5 inch
 - ii. Could be used in Open procedures
 - iii. Applicable for Medium/Large size (green) clips (LT300)
 - iv. Tip angulated 15 degree with facility for distal tip closure and without ratchet
 - v. ETO Sterilizable
 - vi. Can be used in Multiple times (reusable)
 - vii. USFDA Certified

7. Titanium clips medium/large size Pre-Sterilised Ligating clip cartridges, LT300 green:

- One piece set must comprise of minimum 6 such clips.
- It should be compatible with the above mentioned clip applicator and both these items must be from the same manufacturer. (USFDA certified)

8. Open Clip Applicator for clip size - Large (yellow) 7.5 inch:

- It should be:
 - i. Length 7.5 inch
 - ii. Could be used in Open procedures
 - iii. Applicable for Large size (yellow) clips (LT400)
 - iv. Tip angulated 15 degree with facility for distal tip closure and without ratchet
 - v. ETO Sterilizable
 - vi. Can be used in Multiple times (reusable)
 - vii. USFDA Certified

9. Open Clip Applicator for clip size - Large (yellow) 10.5 inch:

- It should be:
 - i. Length 10.5 inch
 - ii. Could be used in Open procedures
 - iii. Applicable for Large size (yellow) clips (LT400)
 - iv. Tip angulated 15 degree with facility for distal tip closure and without ratchet
 - v. ETO Sterilizable
 - vi. Can be used in Multiple times (reusable)
 - vii. USFDA Certified

10. Titanium clips Large size Pre-Sterilised Ligating clip cartridges, LT400 yellow:

- One piece set must comprise of minimum 6 such clips.
- It should be compatible with the above mentioned clip applicator and both these items must be from the same manufacturer. (USFDA certified)

Terms and conditions:

1. All the items should be from the same manufacturing company
2. Item wise selection as L-1 shall not be done
3. The bidder who has not quoted for all the instruments i.e. the entire set shall not be entertained and bid shall be rejected
4. Items must be USFDA, CE, ISO certified.

**Stores Officer (H)
AIIMS, Raipur (C.G)**