



अखिलभारतीयआयुर्विज्ञानसंस्थान, रायपुर(छत्तीसगढ़)  
**All India Institute of Medical Sciences, Raipur (Chhattisgarh)**  
**Tatibandh, GE Road,**  
**Raipur-492 099 (CG)**  
[www.aiimsraipur.edu.in](http://www.aiimsraipur.edu.in)

AIIMS/R/HS/NEPH/1090/166/2024/482

Date: 22/02/2024

**Inviting Quotations for Purchase of Spares for NAV 250 LPM Portable RO (4 items) for Nephrology Department at AIIMS Raipur.**

**NOTICE INVITING QUOTATION**

Sealed quotations are invited from intending registered Stockist / Distributors/traders having GST and relevant documents for Purchase of **Spares for NAV 250 LPM Portable RO (4 items) for Nephrology Department** at AIIMS Raipur. The quotation with copy of certificate of GST & other documents should be submitted to **Stores Office – Hospital, Room No. –329, C C1 Block, Gate No. 1, AIIMS Raipur up to 28/02/2024 before 03:00 pm**. The quotations/Bids will be opened on the same day i.e. on **28/02/2024 at 03:30pm** Details of item are given as under:-

| Sno. | Specification   | Quantity Required | Unit        | HSN Code | Brand | Unit Rate | GST  |      |      | Amount |
|------|---|-------------------|-------------|----------|-------|-----------|------|------|------|--------|
|      | <b>Spares for NAV 250 LPM Portable RO</b>   |                   |             |          |       |           | SGST | IGST | CGST |        |
| 1.   | <b>Activated Carbon Filter</b>  | <b>08</b>         | <b>Pcs.</b> |          |       |           |      |      |      |        |
| 2.   | <b>Softener Cartridge Filter</b>  | <b>08</b>         | <b>Pcs.</b> |          |       |           |      |      |      |        |
| 3.   | <b>UV Lamp</b>  | <b>04</b>         | <b>Pcs.</b> |          |       |           |      |      |      |        |
| 4.   | <b>Micro Filter</b>   | <b>10</b>         | <b>Pcs.</b> |          |       |           |      |      |      |        |
|      | <b>Specification as per Annexure - I</b>  |                   |             |          |       |           |      |      |      |        |
|      | <b>The Bidders must submit the Purchase order copies of the same item supplied to other Central Government institutes or AIIMS.</b> |                   |             |          |       |           |      |      |      |        |
|      | <b>Grand Total</b>  |                   |             |          |       |           |      |      |      |        |

**Terms & Condition**

1. Firm to mention Make/Brand name in their quotation.

2. Taxes, if any (Kindly mention in above table) should be clearly mentioned in the offer.
3. Document relating to registration of firm i.e. GST and relevant document should be submitted along with quotation.
4. Supply should be done within 15 days after Placement of PO.
5. Price should be FOR Destination basis (i.e. concerned department).
6. 100% Payment will be released after certification from concerned department.
- 7. Quotation Name/No. and due date of opening must be mentioned on top of envelopes.**
8. LD @ 0.5% of delayed supply per week or part of week for delay of supply of material subject to maximum up to 10% of delayed supply to be deducted.
9. AIIMS Raipur reserves the right to place order for full or part quantity to one or more firms. The AIIMS, Raipur reserves the right to increase/decrease the number of required quantity.
10. All other terms & condition as per GFR 2017.
11. Material to be delivered at Nephrology Department, AIIMS Raipur.
12. Validity of the quotation should be 90 days from the date of opening.
13. Bidders should submit their email id, contact details with GST registration.
- 14. The Bidders must submit the Purchase order copies of the same item supplied to other central government institutes or AIIMS.**

Sd/-

**Stores Officer (H)**

**AIIMS, Raipur (C.G)**

## **Annexure – I**

### **1. Micron Filter**

- ☐ **Filter Material: Polypropylene (PP)**
- ☐ **Capacity: 1500 LPH**
- ☐ **Filter size: 8.5 micron**
- ☐ **Diameter: 112 mm**

### **2. Activated Carbon Filter**

- ☐ **Filter Material: 1100 iv value**
- ☐ **Capacity: 1500 LPH**
- ☐ **Filter size: 5 micron**
- ☐ **Diameter: 85**

### **3. Softener Cartridge**

- ☐ **Filter Material: Proprietary**
- ☐ **Capacity: 1500 LPH**
- ☐ **Filter size: 20 micron**
- ☐ **Diameter: 70**

### **4. UV Lamp**

- ☐ **16W**
- ☐ **with 118 adapter**