



अखिल भारतीय आयुर्विज्ञान संस्थान, रायपुर (छत्तीसगढ़)
All India Institute of Medical Sciences, Raipur (Chhattisgarh)
Tatibandh, GE Road, Raipur-492 099 (CG)
[Email- storesofficer.hp@aiimsraipur.edu.in](mailto:storesofficer.hp@aiimsraipur.edu.in)

No. AIIMS/R/HS/BME/2025/155/657

Date 29.01.2025

Sub: Inviting quotations for procurement of 5 Lead ECG Cable & NIBP Hose Pipe (for patient monitor) for the department of Biomedical Engineering at AIIMS Raipur.

NOTICE INVITING QUOTATION

Sealed quotations are invited from intending registered Stockiest/Distributors/Traders having GST No./relevant documents for procurement of **5 Lead ECG Cable & NIBP Hose Pipe (for patient monitor) for the department of Biomedical Engineering at AIIMS Raipur.** The quotation with copy of certificate of GST & other documents should be submitted to **office of Stores Officer – Hospital, Room No. – 146, C-C1 Block, Gate No. 1, AIIMS Raipur up to 05/02/2025 before 03:00 pm.** The quotations will be opened on the same day at 03:30pm. Details of items are as under :-

S.n क्र.स.	Name of the Item आइटम कानाम	HSN code एच.एस. एन.	Make/ Model	Maximum Order Qty. मात्रा	UNIT RATE IN Rs. इकाई दररुमें	GST जीएसटी	UNIT RATE with GST इकाई दर करसहित	Total amount in Rs. कुल मूल्य
1.	5 Lead ECG Cable for our existing Mindray (T8, IPM 10, T5) Monitors and accessories			50 nos.				
2.	5 Lead ECG Cable for our existing Philips (MX450, GS 20) Monitors and accessories			150 nos.				
2.	NIBP Hose Pipe for our existing Mindray (T8, IPM 10, T5) Monitors and accessories			50 nos.				
	NIBP Hose Pipe for our existing Philips (MX450, GS 20) Monitors and accessories			100 nos.				
The items should be compatible with our existing patient monitor Specification of the product attached at Annexure-A								

Terms & Condition:-

1. Firm to mention Make/Brand name in their quotation.
2. Taxes, if any (Kindly mention in above table) should be clearly mentioned in the offer.
3. Document relating to registration of firm i.e. GST and relevant document should be submitted along with quotation.

4. Products are certified from ISI/ISO/CE/BIS as applicable, the Certificate to this effect should be attached.
5. Supply should be done within 15 days after Placement of PO.
6. Price should be FOR Destination basis (i.e. concerned department).
7. 100% Payment will be released after certification from concerned department.
8. **Quotation Name/No. and due date of opening must be mentioned on top of envelopes.**
9. LD @ 0.5% of delayed supply per week or part of week for delay of supply of material subject to maximum up to 10% of delayed supply to be deducted.
10. AIIMS Raipur reserves the right to place order for full or part quantity to one or more firms. The AIIMS, Raipur reserves the right to increase/decrease the number of required quantity.
11. All other terms & condition as per GFR 2017.
12. Material to be delivered at Biomedical Engineering department at AIIMS Raipur.
13. Validity of the quotation should be 90 days from the date of opening.
14. Firm to submit documentary evidence in support of their claim for GST at the time of submission of bills.
15. **AIIMS, Raipur reserve the right to ask the bidders for arranging demonstration of their samples for feel & finish for which rates have been quoted to the concerned committee, if required.**

**Officer In-charge
Procurement of DMC
AIIMS Raipur (C.G)**

Annexure-A**Specifications of 5 Lead ECG Cable for patient monitor:**

➤	Category	ECG Cable
➤	Certification	ISO/CE/BIS
➤	Connector Distal	6 Pin Male Round Connector
➤	Connector Proximal	Button/Snap
➤	Latex-free	Yes
➤	Lead Cable Color	Bone Grey
➤	Lead Cable Diameter	3.4 mm
➤	Lead Cable Length	1 Mtr.
➤	Lead Cable Material	PVC
➤	No. of Leads	5
➤	Patient Size	Adult/Pediatric
➤	Total Cable Length	3 Mtr.
➤	Main Cable Color	Grey
➤	Main Cable Length	2 Mtr.
➤	Main Cable Material	PVC
➤	Compatible with	Philips MX450, GS-20, Mindray T8, T5,IPM Series

Specifications of NIBP Hose Pipe for patient monitor:

➤	Certification	ISO/CE/BIS
➤	Connector Distal	Male HP Moulded Connector
➤	Connector Proximal	Push Pull Connector
➤	Hose Color	Grey
➤	Hose Diameter	Inner 3.5 mm x Outer 12.0 mm
➤	Hose Length	3 Mtr.
➤	Hose Type	Single Tube
➤	Latex-free	Yes
➤	Compatible with	Philips MX450, GS-20, Mindray T8, T5,IPM Series

**Officer In-charge
Procurement of DMC
AIIMS Raipur (C.G)**