



आरोग्यं मृतं सम्पदा

अखिल भारतीय आयुर्विज्ञान संस्थान, रायपुर(छत्तीसगढ़)
All India Institute of Medical Sciences, Raipur (Chhattisgarh)
Tatibandh, GE Road,
Raipur-492 099 (CG)
www.aiimsraipur.edu.in

AIIMS/R/HS/2025/155/SPO2N/ 689

Date: 15/04/25

Inviting Quotations for SPO2 PROBE NEONATAL for Neonatology NICU Department AIIMS Raipur.

NOTICE INVITING QUOTATION

Sealed quotations are invited from intending registered Stockist / Distributors/traders having GST and relevant documents for Purchase of SPO2 PROBE NEONATAL for Neonatology NICU Department AIIMS Raipur. The quotation with copy of certificate of GST & other documents should be submitted to office of Hospital Stores Office, FIRST FLOOR Hospital Administrative block Gate No 01, AIIMS Raipur up to 24/04/25 before 03:00 pm. The quotations/Bids will be opened on the same day i.e. on 24/04/25 03:30pm. Details of item are given as under:

Sn o.	Specification	QT Y. RE Q.	Unit	HS N Co de	Brand	Unit Rate	GST			Amount
							SG ST	IGS T	CGST	
1	SPO2 PROBE NEONATAL COMPATIBLE WITH NIHON KOHDEN MONITOR	25	Nos.							
Grand Total										


Terms & Condition

1. Firm to mention Make/Brand name in their quotation.
2. Taxes, if any (Kindly mention in above table) should be clearly mentioned in the offer.
3. Document relating to registration of firm i.e. GST and relevant document should be submitted along with quotation.
4. Products are certified from ISI/ISO/CE/GMP/BIS as applicable, the Certificate to this effect should be attached.
5. Supply should be done within 15 days after Placement of PO.
6. Price should be FOR Destination basis (i.e. concerned department).
7. 100% Payment will be released after certification from concerned department.
8. Quotation Name/No. and due date of opening must be mentioned on top of envelops.
9. LD @ 0.5% of delayed supply per week or part of week for delay of supply of material subject to maximum up to 10% of delayed supply to be deducted.
10. AIIMS Raipur reserves the right to place order for full or part quantity to one or more firms. The AIIMS, Raipur reserves the right to increase/decrease the number of required quantity.
11. All other terms & condition as per GFR 2017.
12. Material to be delivered Neonatology NICU Department AIIMS Raipur.

13. Coding on the packaging "AIIMS Raipur supply not for sale / GOVT. Supply not for sale"
14. Batch testing report submits at the time of supply.
15. Validity of the quotation should be 90 days from the date of opening.
16. Bidders should submit their email id, contact details with GST registration.
17. AIIMS Raipur reserves the right to ask the tenders for arranging demonstration of their samples finish for which rates have been quoted, to the concerned committee, if required.



Office In charge
Procurement of DMC (H)
AIIMS, Raipur (C.G)



Technical Specifications For Monitor Accessories

SPO2 Sensor: - 21 units

1. Neonatal SPO2 sensor should be able to connect with Nihon Kohden monitors in NICU of AIIMS Raipur.
2. The SPO2 Sensor should work on blu-pro / FAST technology. (Documents to be enclosed with technical bid)
3. It should work on the principle of 2 wavelength light absorption method.
4. It should be capable to measure SPO2 in low perfusion. (White Papers confirming the low perfusion technology to be enclosed with technical bid)
5. SPO2 sensor should be water washable for cleaning purpose.
6. It should pass drop test from a height of upto 10 feet.
7. It should be usable at multiple sites with help of a fixing tape.
8. Fixing tape should be provided along with the SPO2 sensor.
9. Sample may be asked for demonstration purpose to verify the above mentioned specifications.