



अखिल भारतीय आयुर्विज्ञान संस्थान, रायपुर(छत्तीसगढ़)
All India Institute of Medical Sciences, Raipur (Chhattisgarh)
Tatibandh, GE Road,
Raipur-492 099 (CG)
www.aiimsraipur.edu.in

AIIMS/R/CS/CON/CHN/59/E-17823

Date: 22/04/2025

QUOTATION NOTICE

Sub: - Request for E-Market Survey for submission of Price Quotation for IUD Insertion Trainer for Dept of College of Nursing at AIIMS Raipur under GFR-2017, Rule155.

Sealed quotations are invited from intending registered Stockist / Distributors having GST and relevant documents for supply of IUD Insertion Trainer for Dept of College of Nursing at AIIMS Raipur. The quotation with copy of certificate of GST & other documents should be submitted to office of **Stores Officer (Central), Medical College building, Gate No-05 at AIIMS Raipur up to 30-04-2025 before 03:00 pm.** The quotations will be opened on the same day at 03:30pm. Details of item are given as under:-

S.no	description	HSN Code	Qty	unit price	GST%	unit price with GST	Total Price
1	IUD Insertion Trainer		08				
	Total		08				

Terms and condition

- 1) Price Quotation submission: The price Quotation must be submitted in the sealed envelope containing all requested documents/sample on the Letter Head of the Firm which should contain a valid GST Registration number clearly, and shall be dropped in the Tender box kept in the Central Stores ,Ground floor, Medical college building, AIIMS Raipur before the stipulated due date for submission mentioned in this NIQ.
- 2) Rate should be mentioned in words & figure both.
- 3) Delivery Schedule – within 21 days from the date of issue of PO.
- 4) Price should be FOR Destination basis. (i.e Department of College of Nursing, AIIMS Raipur).
- 5) LD @0.5% of delayed supply per week or part of week for delay of supply of material subject to maximum up to 10%.
- 6) After expiry of delivery Period material cannot be accepted without the extension of delivery period.
- 7) Quotation No/Name and Due date of opening must be written on top of envelop.
- 8) GST rates applicable on your quoted item may please be confirmed. HSN code for each item shown be clearly mentioned.
- 9) Please confirm if there any change (Upward/Reduction) in your Basic Price structure. And you are also requested to pass the Input Credit as per the following Anti Profiteering Clause of GST **“Upon Implementation of GST, any reduction in the rate of tax on supply of goods or service or the benefit of input tax credit shall be passed on to AIIMS Raipur by way of commensurate reduction in the prices”.**
- 10) The GST registration details may please be furnished.
- 11) 100% payment against receipt and acceptance of material.

- 12) Validity of offer should not be less than 90 days.
- 13) No part supply or part payment will be entertained.
- 14) Expenditure will be debitable to GIA-GENERAL.
- 15) Brand, Make & warranty should be clearly mentioned in offer (if required)
- 16) The Quantity of above column is totally tentative. It can be increased or decreased at the time of placement of order.
- 17) Supply, installation & commissioning will be done by Firm.(if applicable)
- 18) The product should be of superior quality and highly standard.

General Technical Specifications for IUD Insertion Trainer:

1. Material: The trainer should be made of soft, lifelike materials, such as silicone or rubber, that closely mimic postpartum human uterine tissue.
2. Anatomical Accuracy: It must accurately represent female pelvic anatomy, including the cervix, uterus, and fallopian tubes.
3. Functional Capabilities: The uterus should facilitate the insertion of an IUD and uterine balloon tamponade; it should also allow for vaginal examinations and enable learning of both anteverted and retroverted positions of the uterus.
4. Realistic Insertion Experience: The trainer should allow for simulated dilation and provide realistic resistance during IUD insertion.
5. Insertion Techniques: The model should demonstrate both manual and instrument-assisted techniques for IUD insertion.
6. Correct IUD Placement: The design must support the correct positioning and placement of the IUD.
7. Removable IUD: It should include a removable IUD for practicing both insertion and removal.

8. Tools Included: The trainer should be supplied with essential instruments, including PPIUCD forceps, sponge-holding forceps, Sims speculum, vulsellum forceps, Cusco's speculum, uterine sound, MVA cannula, and artery forceps.

9. Maintenance: Replaceable parts should be included for easy cleaning and repair.



**Stores Officer
AIIMS RAIPUR**

