



No.: AIIMS/R/CS/PATHO/14-48/E-19547/25/LPC

Date: 21/05/2025

(QUOTATION NOTICE)

Request for E-Market Survey for submission of Price Quotation for “Variable Volume Single Channel Micropipette for Pathology & Lab Medicine at Department at AIIMS, Raipur under GFR-2017, Rule-155.

NOTICE INVITING PRICE QUOTATION UNDER GFR-2017, RULE-155

Sealed quotations are invited from intending registered Stockiest/ Distributors having GST/ relevant documents for above said. The quotation should be submitted to Medical College Building, Ground Floor, Gate No. 05, AIIMS, Raipur office of the Stores Officer up to 27/05/2025 before 3:00 pm. The quotations will be opened on the same day at 03:30 pm. detailed specifications of items are as under:

| Sr. No. | Item of Description | Qty. | Make/ Model | HSN Code | Unit Price (Rs.) | GST | Unit Rate with GST | Total Price |
|---------|------------------------------------------------------------------------------------------------------------------------|------|-------------|----------|------------------|-----|--------------------|-------------|
| 01 | Variable Volume Single Channel Micropipette (0.5-10ul). Accuracy: 0.1ul Precision: 0.05ul Subdivision: 0.01ul | 03 | | | | | | |
| 02 | Variable Volume Single Channel Micropipette (02-20ul). Accuracy: 0.16ul Precision: 0.08ul Subdivision: 0.02ul | 03 | | | | | | |
| 03 | Variable Volume Single Channel Micropipette (05-50ul). Accuracy: 0.4ul Precision: 0.15ul Subdivision: 0.05ul | 06 | | | | | | |
| 04 | Variable Volume Single Channel Micropipette (10-100ul). Accuracy: 0.6ul Precision: 0.02ul Subdivision: 0.1ul | 04 | | | | | | |
| 05 | Variable Volume Single Channel Micropipette (20-200ul). Accuracy: 1.2ul Precision: 0.4ul Subdivision: 0.21ul | 06 | | | | | | |
| 06 | Variable Volume Single Channel Micropipette (100-1000ul). Accuracy: 6.0ul Precision: 2.0ul Subdivision: 1.0ul | 07 | | | | | | |

Terms and Condition

1. Rate should be mentioned in words & figure both.
2. Taxes/GST, if any (Kindly mention in above table)
3. Delivery schedule: within 04 Weeks from the date of issue of P.O. Delivery period:
4. No additional documents related to this NIQ will be entertained after opening of NIQ.
5. Price should be FOR Destination basis (i.e. Department of Patho and Lab Medicine, AIIMS Raipur).
6. LD @ 0.5% of delayed supply per week or part thereof for delay of supply of material subject to maximum up to 10% should be deducted. After expiry of delivery period material cannot be accepted without extension of delivery period.
7. **The item should be specified as per ANNEXURE-01 (Technical Specification).**
8. Quotation No/Name and due date of opening must be written on top of envelop.
9. GST rates applicable on your quoted item may please be confirmed. **HSN code for item should be clearly mentioned.**
10. Any change (Upward/Reduction) in the Basic Price structure post submission of bid shall be promptly intimated to AIIMS Raipur. You are also required to pass the Input Credit as per the following **Anti Profiteering Clause of GST.**
“Upon Implementation of GST, any reduction in the rate of tax on supply of goods or service or the benefit of input tax credit shall be passed on to AIIMS Raipur by way of commensurate reduction in the prices”.
11. The GST registration details may be furnished.
12. RTGS details required for payment purpose.
13. 100% payment against receipt and acceptance of material.
14. No part payment will be entertained.
15. Validity of offer should not be less than 90 days.
16. Supply, installation and commissioning will be done by firm (if applicable).
17. The quantity in above column is totally tentative. It can be increased or decreased at the time of placement of order.
18. AIIMS, Raipur reserve the right to place the order for full or part quantity to one or more firms.
19. The product should be of superior quality and highly standard.
20. **L-1 (Only technically qualified by department) of items will be calculated on total value of item.**


Stores Officer,
AIIMS, Raipur (C.G.) (क्रय)
Stores Officer (CP)
एम्स, रायपुर (छ.ग.)
AIIMS Raipur (C.G.)

Technical Specifications of Variable Volume Single Channel Micropipette.

1. Pipette should be light weight with high precision, robust and dependable.
2. Pipette should have effortless one hand operation like Volume setting, Volume locking, Pipetting and tip ejectio for both right and left handlers.
3. Pipette should have 4 position volume displays with magnifying lens.
4. Display always visible and facing the user during operation.
5. The Pipette should have volume locking provision to avoid accidental volume range.
6. All operations and operating keys should be easily accessible with thumb of operating hand from one position for operator's convenience and to minimise strain on finger.
7. Short operating stroke length of the pipetting key max 12 mm at nominal volume.
8. Color code for volume and tips to be used should be clearly visible at least at 3 points.
9. Slim shafts to reach narrow plastic ware like centrifuge tubes.
10. Easy to dismantle & reassemble without any tools, all parts should be replaceable.
11. The pipette should be fully autoclavable without disassembly-for maximum protection from contamination.
12. The pipette should have corrosion resistant piston for volume range 0.1ul to 20ul (Stainless Steel); 10ul to 1000ul (PEEK).
13. The pipette should have a warranty of 3 years.
14. Individual labelling area for user specific identification.
15. Easy Calibration Technique so it enables user to calibrate and adjust the instrument in house without any tools.
16. UV Resistant.
17. CE Compliant.
18. Onsite demonstration is required at the time of Technical Evaluation.

