



No: AIIMS/R/CS/URO/2025/59/ST

Date: 28.05.2025

Call for Objection

Subject: Inviting comments/objection, if any before declaring proprietary article for procurement of "Flexible Digital Ureterorenoscope (Compatible with already installed Biorad Medisys Video processor for Urology at AIIMS, Raipur.


Urology Department at AIIMS, Raipur has to procure Flexible Digital Ureterorenoscope through Proprietary Article basis.

The proposal submitted by M/s Biorad Medisys Pvt. Ltd, Maharashtra, India who is sole manufactures and M/s Sai Associates, Durg is the Local Agent of this item along with Proprietary Article Certificate are attached & uploaded on CPP Portal & Institute website.

The above documents are being uploaded for open in formation to submit objections, comments if any from any manufacturer/supplier before declaring proprietary article of the said equipment/items to be procured, within 10 days (i.e. 07.06.2025) from the date of issuance/uploading of the notification.

The objection should be raised in the technical compliance sheet as enclosed, if any firm claiming suitability of their product with respect to specification mentioned.

The comments should be sent to the office of Central Store Office on above address at AIIMS Raipur in a sealed envelope with above reference on sent to respective mail address storesofficer.cp@aiimsraipur.edu.in on before 07.06.2025 up to 05:00PM from the date of uploading on institutional website, failing which it will be presumed that any other manufacture/vendor is having no comment to offer and case will be decided on merits.


Stores Officer
AIIMS Raipur (CG)
भारत अधिकारी (के क्रय)
Stores Officer (CP)
एम्स, रायपुर (छ.ग.)
AIIMS Raipur (C.G.)

Enclosure:-

1. PAC Certificate by Manufacturer.
2. PAC Certificate by Department.

Annexure - I

Specifications for Flexible Digital Ureterorenoscope (Compatible with already installed Biorad Medisys Video Processor Unit Model: SD300A)

1.1) Digital CMOS imager, Chip on Tip technology.:

1. Flexible Digital Ureterorenoscope

- 1.2) Direction of view: zero degree (forward-viewing)
- 1.3) Field of view: 110-degrees or more
- 1.4) Observation depth of field: 2mm–50mm
- 1.5) Integrated camera head; no secondary external attachments required.
- 1.6) Can have limit of the number of hours of usage for each scope, with at least 15 hours. Usage time counter of each scope must be displayed on screen from beginning.
- 1.7) Flexible insertion sheath with logical deflection mechanism.
- 1.8) Active deflection tip length should be 6.5 cm or more.
- 1.9) Deflection of scope tip should be possible to 270-degrees in both the directions.
- 1.10) Insertion tip diameter: 7.7F or less.
- 1.11) Insertion sheath outer diameter: 9.9F or less.
- 1.12) Working channel [ID]: 3.3F or more.
- 1.13) Working length: 650 to 670mm
- 1.14) Compatible with laser lithotripsy: uses existing technologies and familiar surgical tools.

2. It must be compatible with already installed Biorad Medisys Video Processor Unit Model: SD300A at the department.

3. Reprocessing method compatibility:

- 3.1) It must be compatible for sterilization method with ETO, Steris & Sterrad.
- 3.2) It should also be compatible for High Level Disinfection by immersion in liquid chemicals like Cidex & Paracef.

4. All instruments should be CESDO India/ European CE with 4-digit notified body no./ US FDA approved for the intended use.



अखिल भारतीय आयुर्विज्ञान संस्थान, रायपुर (छ.ग.)
All India Institute of Medical Sciences, Raipur (Chhattisgarh)
Tatibandh, GE Road,
Raipur-492 099 (CG)
www.aiimsraipur.edu.in

स्वामित्व प्रमाण पत्र
Proprietary Article Certificate

फाइल संख्या और संदर्भ File Number and Reference		No./Urology/2025/OW/59
1	सामाग्री का विवरण Description of article	FLEXIBLE DIGITAL URETERO-RENSCOPE
2	पूर्वानुमानित मात्रा/वार्षिक आवश्यकता Forecast of quantity/annual requirement	10 (TEN)
3	उपरोक्त मात्रा हेतु अनुमानित मूल्य Approximate estimated value for above	[REDACTED]
4	निर्माता का नाम एवं पता Maker's name and address	BIORAD MEDISYS PVT. LTD., INDIA
5	अधिकृत डीलर/स्टाकिस्ट का नाम Name(s) of authorised dealers/stockists	
6	<p>मैं पी ए सी के आधार पर उपरोक्त खरीद को स्वीकार करता हूं और यह प्रमाणित करता हूं कि:</p> <p>नोट- (बी), (सी-1) या (सी-2) में से केवल एक को बनाए रखने के लिए टिक करें, जो भी लागू हो और दूसरो को काट दें। कृपया (ए) टिक कर पुष्टि करें इसके बिना पीएसी प्रमाण पत्र अवैध होगा</p> <p>I approve the above purchase on PAC basis and certify that:-</p> <p>Note- Tick to retain only one out of (b), (c-1) or (c-2) whichever is applicable and cross out others. Please do confirm (a) by ticking it – without which PAC certificate will be invalid.</p>	
6 (a)	यह एकमात्र फर्म है जो इस मद का निर्माण/संग्रहण कर रहा है। और This is the only firm who is manufacturing /stocking this item. AND	✓
6 (b)	किसी अन्य फर्म द्वारा समरूप मद निर्मित/विक्रय नहीं किया जाता है, जिसका उपयोग इसके बदले किया जा सकता है। अथवा A similar article in not manufacturing/sold by any other firm, which could be used in lieu OR	
6 (c-1)	कोई अन्य मेक/ब्रांड निम्नलिखित कारणों (जैसे ओईएम/वारंटी के) के लिए उपयुक्त नहीं होगा। अथवा No other make/brand will be suitable for following tangible reasons (like OEM/warranty spares): OR	✓
6 (c)	कोई अन्य मेक/ब्रांड निम्नलिखित कारणों से उपयुक्त नहीं होगा (अगर पीएसी	

	<p>पिछले खरीद में भी दिया गया था, तो कृपया इसके बाद से अधिक स्रोतों का पता लगाने के लिए प्रयास करें) तथा</p> <p>No other make/brand will be suitable for following intangible reasons (if PAC was also given in the last procurement cycle, please also bring out efforts made since then to locate more sources): OR</p> <p>.....</p> <p>.....</p>	
7	<p>प्रस्ताव के लिए वित्त शाखा की सहमति का संदर्भ (कार्रवाई भंडार और लेखा विभाग द्वारा की जायेगी)</p> <p>Reference of concurrence of finance wing to the proposal (Action will be taken by stores & Account Department)</p>	<p>.....</p> <p>.....</p>

पिछले तीन सालों में इस मद की पीएसी खरीद का इतिहास नीचे दिया जा सकता है (यदि कोई हो) History of PAC purchase of this item for past three years may be given below (if any)

प्रदायक का नाम Name of the Supplier			
आदेश/निविदा संदर्भ और दिनांक Order/Tender reference & Date	आदेशित मात्रा Quantity Ordered	आदेश पर मूल दर (₹) Basic Rate on order (Rs.)	प्रतिकूल प्रदर्शन रिपोर्ट अगर कोई हो Adverse Performance Reported if any

अनुमोदन करने वाले प्राधिकारी का हस्ताक्षर -----

Dr. Deepak Biswal

दिनांक -----

22-3-2025

DR. DEEPAK BISWAL

अधिकारी का पदनाम -----