

अखिल भारतीय आयुर्विज्ञान संस्थान,रायपुर(छत्तीसगढ़)

All India Institute of Medical Sciences, Raipur (Chhattisgarh)

Tatibandh, GE Road, Raipur-492 099 (CG)

www.aiimsraipur.edu.in

Date: 20.06.2025

No.: AIIMS/R/CS/ENTSMP/25/LPC

(QUOTATION NOTICE) 1st call

Request for E-Market Survey for submission of Price Quotation for "Supply of Eustachian Tube Balloon Dilatation kit" for ENT Department at AIIMS, Raipur under GFR-2017, Rule-155 For use in Project

NOTICE INVITING PRICE QUOTATION UNDER GFR-2017, RULE-155

Sealed quotations are invited from intending registered Stockiest/ Distributors having GST/ relevant documents for above said. The quotation should be submitted to Department of ENT & Head Neck Surgery OPD, ROOM-12, C-block Ground Floor, Gate No. 04, AIIMS, Raipur up to 30.06.2025 before 2:00 pm. The quotations will be opened on the same day at 03:00 pm. Detailed specifications of items are as under:

Eustachian Tube Balloon Dilatation Catheter- It should have a non complaint Balloon with puncture resistance. It should have a working length of 256mm. It should have a flexible shaft. It should have marker bands for endoscopic & direct visualization. It should have arrive port for inflation / deflation of the guide wire. Balloon should reach its stated size at 6mm * 14mm Atmosphere and should withstand 10 atmospheric pressure- 6mm x 14 mm length. It should contain 55 degree guide wire Eustachian Tube Balloon Dilatation Catheter- It should have a non complaint Balloon with puncture resistance. It should have a working length of 256mm. It should have a flexible shaft. It should have marker bands for endoscopic & direct visualization. It should have separate port for inflation / deflation and a separate port for infroduction of the guide wire. Balloon should reach its stated size at 7mm* 14mm Atmosphere and should withstand 8 atmospheric pressure- 7mm x 14 mm length. It should contain 55 degree quidewire	Sl. No	Item of Description	Qt y.	Make/ Model	HSN Code	Unit Price (₹)	GST	Unit Rate with GST	Total Price
Dilatation Catheter- It should have a non complaint Balloon with puncture resistance. It should have a working length of 256mm. It should have a flexible shaft. It should have marker bands for endoscopic & direct visualization. It should have separate port for inflation / deflation and a separate port for inflation should reach its stated size at 7mm* 14mm Atmosphere and should withstand 8 atmospheric pressure 7mm x 14 mm length. It should contain 55	01	Catheter- It should have a non complaint Balloon with puncture resistance. It should have a working length of 256mm. It should have a flexible shaft. It should have marker bands for endoscopic & direct visualization. It should have separate port for inflation / deflation and a separate port for introduction of the guide wire. Balloon should reach its stated size at 6mm * 14mm Atmosphere and should withstand 10 atmospheric pressure 6mm x 14 mm length. It should contain 55 degree guide	3						
	02	Eustachian Tube Balloon Dilatation Catheter- It should have a non complaint Balloon with puncture resistance. It should have a working length of 256mm. It should have a flexible shaft. It should have marker bands for endoscopic & direct visualization. It should have separate port for inflation / deflation and a separate port for introduction of the guide wire. Balloon should reach its stated size at 7mm* 14mm Atmosphere and should withstand 8 atmospheric pressure 7mm x	3						

Terms and Condition

- 1. Rate should be mentioned in words & figure both.
- 2. Taxes/GST, if any (Kindly mention in above table).
- 3. Delivery schedule: within 02 Weeks from the date of issue of P.O. Delivery period:
- 4. No additional documents related to this NIQ will be entertained after opening of NIQ.
- 5. Price should be FOR Destination basis (i.e. Department of ENT), AIIMS Raipur).
- 6. LD @ 0.5% of delayed supply per week or part thereof for delay of supply of material subject to maximum up to 10% should be deducted. After expiry of delivery period material cannot be accepted without extension of delivery period.
- 7. Quotation No/Name and due date of opening must be written on top of envelop.
- 8. GST rates applicable on your quoted item may please be confirmed. **HSN code for item should be clearly mentioned**.
- 9. Any change (Upward/Reduction) in the Basic Price structure post submission of bid shall be promptly intimated to AIIMS Raipur. You are also required to pass the Input Credit as per the following **Anti Profiteering Clause of GST**.
 - "Upon Implementation of GST, any reduction in the rate of tax on supply of goods or service or the benefit of input tax credit shall be passed on to AIIMS Raipur by way of commensurate reduction in the prices".
- 10. The GST registration details may be furnished.
- 11. RTGS details required for payment purpose.
- 12. 100% payment against receipt and acceptance of material.
- 13. No part payment will be entertained.
- 14. Validity of offer should not be less than 90 days.
- 15. Supply, installation and commissioning will be done by firm (if applicable).
- 16. The quantity in above column is totally tentative. It can be increased or decreased at the time of placement of order.
- 17. AIIMS, Raipur reserve the right to place the order for full or part quantity to one or more firms.
- 18. The product should be of superior quality and highly standard.
- 19. L-1 of items will be calculated on total value of all items.

Dr. Sharmistha Chakravarty
Associate Professor,
Principal Investigator,
Department of ENT & Head Neck Surgery
AIIMS, Raipur,
Chhattisgarh