



अखिलभारतीयआयुर्विज्ञानसंस्थान, रायपुर (छत्तीसगढ़)  
All India Institute of Medical Sciences, Raipur (Chhattisgarh)  
Tatibandh,  
Raipur-492 099 (CG)

AIIMS/R/HS /Micro/2025/212/LPC/B/E-23627/ 856

Date: 30/06/2026

विषय/Sub: Inviting Quotation for two bid system (Technical and Financial) for the procurement of "Consumable Items(2<sup>nd</sup> Call)" at Microbiology Department of AIIMS Raipur.

### कोटेशनसूचना

#### QUOTATION NOTICE

Two separate sealed quotation is invited from intending registered stocklist/ Distributors having GST Number and relevant documents for "Consumables Items (2<sup>nd</sup> Call)" for Microbiology Department, AIIMS Raipur. The quotation should be submitted to office of Stores Officer (Hospital), Hospital Administrative block, Old MS Office, Gate No. 1, AIIMS Raipur up to 03:00 pm on 06/07/2026. The quotations will be opened on the same day at 3:30 pm. Details of item are given as under:-

Sr. No.	NAME	REQUIREMENT	UNIT	Make & HSN Code	UNIT RATE IN Rs.	GST in %	Unit Rate with GST	TOTAL PRICE in RS.
1	Echinococcus IgG Elisa Kit	(1x 192) x 5	Kit					
2	Entamoeba Histolytica IgG Elisa Kit	( 1x 96) x 2	Kit					

#### Terms & Conditions.

1. The supplier is required to supply the kits in staggered lots.
2. The annual requirement for the specified kits is as follows: 5 kits of Echinococcus IgG ELISA Kit and 2 kits of Entamoeba Histolytica IgG ELISA Kit. As per the delivery schedule, 3 kits of Echinococcus IgG ELISA Kit are to be delivered in the first 6 months and 2 kits in the subsequent 6 months; 1 kit of Entamoeba Histolytica IgG ELISA Kit is to be delivered every 6 months.
3. Interested agencies must submit sealed technical bids and sealed financial bids separately. Sealed Cover-I should contain the "Technical Bid" and Sealed Cover-II should contain the "Financial Bid." Both these covers must be placed inside a third sealed cover bearing all necessary details (the quotation number/name and opening date must be written on the envelope).
4. The technical bid must contain complete details of the items and other relevant documents requested in the NIQ.
5. If required, samples will be sought from bidders found to be technically qualified.
6. Financial bids will be opened only for those bidders whose samples are found to be technically acceptable.
7. Rates must be quoted in both words and figures.
8. The applicable GST rate on the quoted items must be specified.
9. Delivery Schedule: Within 15 days from the date of issue of the Purchase Order (P.O.).
10. No additional documents related to this NIQ will be accepted after the NIQ opening.
11. Prices must be quoted on an F.O.R. basis (delivery up to the destination/concerned department).

12. In case of a delay in the supply of material, Liquidated Damages (LD) at the rate of 0.5% of the value of the delayed supply per week (or part thereof) of delay will be deducted, subject to a maximum of 10% of the value of the delayed supply. 14. Submission of the GST registration certificate is mandatory.
15. Submission of the concerned firm's PAN card & GST registration certificate are also mandatory.
16. It is mandatory for the participating firm to submit its drug license as well.

  
30/06/2026

Stores Officer-H  
AIIMS Raipur (C.G.)

पवन कुमार

**Pawan Kumar**

भंडार अधिकारी (अस्प.)/Stores Officer (Hosp.)  
अखिल भारतीय आयुर्विज्ञान संस्थान, रायपुर (छ.ग.)  
All India Institute of Medical Sciences, Raipur (C.G.)

Technical specification of *Echinococcus* IgG ELISA kit-

1. The ELISA kit should be used for qualitative detection of *Echinococcus granulosus* IgG antibodies in human serum.
  2. Sample type – serum.
  3. No. of wells per kit- 1 x 96 wells or 1x192 wells
  4. The quoted kit must be based upon sandwich ELISA technique.
  5. The microplate wells should be precoated with *Echinococcus cyst* antigen.
  6. Detection method- Chromogenic
  7. Storage- 2°C to 8°C
  8. Sample quantity required- approx. 100 µl
  9. Assay time- ≤ 90 min
  10. The ELISA kit must be highly specific & sensitive.
  11. The kits should be IVD approved
  12. The kit must have long shelf life.
- 

Technical specification of *Entamoeba histolytica* IgG ELISA kit-

1. The ELISA kit should be used for qualitative detection of *Entamoeba histolytica* IgG antibodies in human serum/ plasma.
2. Sample type - serum or plasma.
3. No. of wells per kit- 1 x 96 wells or 1x192 wells
4. The quoted kit must be based upon sandwich ELISA technique.
5. The microplate wells should be precoated with *Entamoeba histolytica* antigen.
6. Detection method- Bichromatic
7. Storage- 2°C to 8°C
8. Sample quantity required- approx 100 µl
9. Assay time- ≤ 125 min
10. The ELISA kit must be highly specific & sensitive.
11. The kits should be IVD approved
12. The kit must have long shelf life.