



अखिलभारतीयआयुर्विज्ञानसंस्थान, रायपुर(छत्तीसगढ़)
All India Institute of Medical Sciences, Raipur (Chhattisgarh)
Tatibandh, GE Road,
Raipur-492 099 (CG)
www.aiimsraipur.edu.in

AIIMS/R/HS/T&E/LPC/155/2026/E-24167 / 868

Date: 20/07/2026

Inviting Quotations for Purchase of **Battery for automated CPR Machine**, for Department **Trauma & Emergency** at AIIMS Raipur.

NOTICE INVITING QUOTATION

A sealed quotation is invited from Authorized Distributors having GST and relevant documents for Purchase of **Battery for automated CPR Machine** for Department of **Trauma & Emergency** at AIIMS Raipur. The quotation with copy of certificate of GST & other related documents should be submitted to Stores Office (Hospital), new administrative building, First Floor, A- Block, Gate no 05, AIIMS Raipur, **up to 27/07/2026 before 03:00 pm**. The quotations/Bids will be opened on the same day i.e. on **20/07/2026 at 03:30pm**.

Details of item are given as under:-

Sn o.	Specification	Quant ity Requi red	Unit	HSN Code	Make	Unit Rate	GST			Amount
							SGS T	IG ST	CG ST	
1.	Battery for automated CPR Machine	04	Nos.							
Specification as per annexure- A										
Grand Total										

Terms & Condition

1. Firm to mention Make/Brand name in their quotation.
2. Taxes, if any (Kindly mention in above table) should be clearly mentioned in the offer.
3. Document relating to registration of firm i.e. GST and relevant document should be submitted along with quotation.

4. Products are certified from ISI/ISO/CE/GMP/BIS as applicable, the Certificate to this effect should be attached.
5. Supply should be done as per Purchase Order.
6. Price should be FOR Destination basis (i.e. concerned department).
7. 100% Payment will be released after certification from concerned department.
8. **Quotation Name/No. and due date of opening must be mentioned on top of envelops.**
9. LD @ 0.5% of delayed supply per week or part of week for delay of supply of material subject to maximum up to 10% of delayed supply to be deducted.
10. AIIMS Raipur reserves the right to place order for full or part quantity
11. The AIIMS, Raipur reserves the right to increase/decrease the number of required quantity.
12. All other terms & condition as per GFR 2017.
13. **Material to be delivered at Department of Trauma & Emergency, AIIMS Raipur.**
14. **Validity of the quotation should be 90 days from the date of opening.**
15. **Bidders should submit their email id, contact details with GST registration.**
16. **AIIMS Raipur reserves the right to ask the tenderers for arranging demonstration of their samples for feel & amp; finish for which rates have been quoted, to the concerned committee, if required.**


Stores Officer
Hospital store
AIIMS, Raipur (C.G)
Pawan Kumar
मंडार अधिकारी (अस्प.)/Stores Officer (Hosp.)
अखिल भारतीय आयुर्विज्ञान संस्थान, रायपुर (छ.ग.)
All India Institute of Medical Sciences, Raipur (C.G.)

Annexure -A

Specification for Rechargeable Battery Compatible with Automated CPR Device (Easy Pulse)

1. Should be compatible with Easy pulse Automated CPR Device (Schiller)
2. **Battery Type:** Rechargeable Lithium-Ion / Lithium-Polymer Battery.
3. **Nominal Voltage:** 18 V to 19 V.
4. **Capacity:** Minimum 4.0 Ah.
5. **Energy Rating:** Minimum 70 Wh.
6. **Cell Configuration:** Multi-cell rechargeable battery pack suitable for medical equipment operation.
7. **Operating Temperature Range:** -20°C to +40°C or better.
8. **Charging Time:**
 - a. Battery should achieve full charge within **75 minutes or less** using the OEM-recommended charger.
 - b. Charging system should support simultaneous charging of multiple batteries, if applicable.
9. **Battery Management System (BMS):**
 - a. In-built protection against overcharging, deep discharge, short circuit, overheating, and overcurrent.
10. **Compatibility:**
 - a. Battery shall be fully compatible with the offered Automated CPR Device and supplied charging station.
11. **Certifications:**
 - a. Battery shall comply with applicable IEC/IS safety standards for rechargeable medical device batteries.
12. **Shelf Life and Service Life:**
 - a. Minimum shelf life of 12 months from date of supply.
 - b. Battery should retain at least 80% of original capacity after 300 charge-discharge cycles or as per OEM specification.
13. **Supply:**
 - a. Battery shall be supplied along with necessary connectors and charging accessories, if required for operation.
14. **Warranty:**
 - b. Minimum 1-year warranty from date of installation/commissioning


22/07/2026
Stores Officer
Hospital store
AIIMS, Raipur (C.G)
Pawan K.
भंडार अधिकारी (अस्प.)/Stores
मुख्य भारतीय आयुर्विज्ञान संस्थान
All India Institute of Medical Sciences, Raipur, C.G.