



No. AIIMS/R/CS/4007/PHYPMR/GTE/

अखिल भारतीय आयुर्विज्ञान संस्थान (रायपुर) छत्तीसगढ़
(All India Institute of Medical Sciences, Raipur (Chhattisgarh))
Tatibandh, GE Road, Raipur-492 099 (CG)
www.aiimsraipur.edu.in

Date - 01/05/2023

Corrigendum

Sub: Corrigendum in Notice Inviting Tender for “Electro Physiology (EP) System” for the department of Physiology and PMR, AIIMS Raipur.

Tender Id: - 2023_IMSRP_747249_1 date 28/03/2023.

In context to following corrigendum is issued against subject:

S.no.	Page/Clause/Point	Existing	To be read as/Addenda
1.	Section –Vii Point -1 Subpoint - a	System should be 08 channels EMG, NCV & EP system with individual touch proof & DIN connectors	System should have at least 8 channels EMG, NCV & EP system with individual touch proof & DIN connectors
2	Section –Vii Point -1 Subpoint c	The amplifier should have USB powered and enabling the amplifier to be connected to any PC/Laptop and transfer data through same USB.	The amplifier should be USB powered/AC powered and enabling the amplifier to be connected to any PC/Laptop and transfer data through same USB.
3	Section –Vii Point -1 Subpoint c	Tremor analysis	Desirable to have Tremor analysis
4	Section –Vii Point -1 Subpoint f	The system should have wireless control panel for easy operation.	The system should have wireless/wired control panel for easy operation.
5	Section –Vii Point - 4 Subpoint c	Supplied with HD-USB camera and software with facility to record Video of the patient while doing any test specially Pattern VEP, this option is useful to watch a patient’s eyes focusing on the fixation point, later on video frame can be copied to an exam report, to an additional information window or to a file.	The system is desirable / preferable to be supplied with HD-USB camera and software with facility to record Video of the patient while doing any test specially Pattern VEP, this option is useful to watch a patient’s eyes focusing on the fixation point, later on video frame can be copied to an exam report, to an additional information window or to a file
6	Section –Vii Point - 4 Subpoint d	The Video Capture Module for Lab Chart is used for simultaneous recording and synchronization of a digital video and a Lab Chart data file.	The system is desirable / preferable to be supplied with the Video Capture Module for Lab Chart is used for simultaneous recording and synchronization of a digital video and a Lab Chart data file.
7	Section –Vii Point - 4 Subpoint m	Dedicated control panel can function via Bluetooth & USB interface both.	Dedicated control panel can function via Bluetooth &/or USB interface
8	Section –Vii Point - 5 Subpoint j	Gold plated cup electrode – 20 pcs.	Gold plated/ silver plated cup electrode – 20 pcs.
9	Point - 15	The warranty shall remain valid for 60 months commencing from first patient treated as per AERB norms	The warranty shall remain valid for 60 months commencing from first patient.

Note :-

*All other terms & condition will be remain unchanged.
Corrigendum must be uploaded along with technical bid as acceptance.*

**Sr.Procurement cum Store Officer
AIIMS Raipur**