



अखिल भारतीय आयुर्विज्ञान संस्थान, रायपुर (छत्तीसगढ़)
All India Institute of Medical Sciences, Raipur (Chhattisgarh)
Tatibandh, GE Road,
Raipur-492 099 (CG)
www.aiimsraipur.edu.ir

AIIMS/R/HS/CSSD/2024/RC/1/2211

Date: 01/01/2025

Corrigendum


Tender Id No.: 2024_IMSRP_840268_1 **Dated:** 18/12/2024.

With reference to above tender ID No., the following amendment is being issued for the tender **“Annual Rate Contract for procurement of Consumables of Detergent, Disinfectant & Lubricant Solution for CSSD departments”** at AIIMS Raipur:-

| SN | Page/ Clause/ Point | Existing Parameters | To be read as |
|----|---|--|---|
| 1. | Description s/Specification Annexure 1 (Page No.15) | 1.Enzymatic Solution 2.Lubricant Spray 3.Disinfectant Solution 4.Acidity Solution | <p>1.Detergent Solution :- (Enzymatic) Detergent should have low foam, multi enzyme solution, and mild alkaline. PH= 9-10.5 Should work in hard, soft and cold water, biodegradable, Eco friendly, No Eco toxicity, CE Certified. Unit Pack size is 1 Litre.</p> <p>2. Instrument Lubricant Solution:- Solution must non- sticky, non toxic silicon free and water soluble mineral oil lubricant. Prevent rust, staining and spotting. Steam penetrable, steam and EO Sterilizable No Eco toxicity, bop degradable. Composite of Butane/Propane <10. PH is 7.9 to 9.6. Certified by ISO and CE, Not flammable, milky white appearances. Unit Pack size is 1 Litre.</p> <p>3. Disinfectant Solution :- It should have Bactericidal, Virucidal, and Fungicidal activity. CE and ISO certified, EPA registered. It must contain n-alkyl dimethyl Benzyl Ammonium Chloride, N-alkyldimethyle, Ethyl benzyl Ammonium Chloride. PH is 11-12. No Eco toxicity must be phosphate free. Unit Pack size is 1 Litre.</p> <p>4. Acidic Detergent Solution:- It should Mild Acidic Solution, PH>1.5. Non corrosive, Restores articulation to box-locks and joints, water soluble. Should be safe efficient and quick revitalizing solution for stainless steel Surgical instrument, tray, and basin and case cart. Unit Pack size is 1 Litre.</p> <p>Notes:- 1. As per the requirement of machine the litres should be converted in Gallons. (I.e. 1 gallon = 4 Litres)</p> |

All the other terms and condition will be remained unchanged.

Dr. Yogendra N. Keche
(Officer In-charge, Procurement of
Drugs, Medicine & Consumable)
AIIMS, Raipur (C.G.)


**Officer Incharge
Procurement (DMC)
AIIMS Raipur (C.G.)**