

Technical Corrigendum For Tender No. AIIMS/R/CS/GENMEDI/ICU/GTE/25

Flexible Video Bronchoscope

S No.	Existing Specification	Revised Specifications
A	HD Video Processor with built in Xenon light source (150Wto 300W).	HD Video Processor with built in Xenon light source (150Wto 300W) / LED Light source having atleast 4 LEDs.
3	It should have xenon short arc lamp of 150W-300Wwith 3W LED as auxiliary lamp	It should have xenon short arc lamp of 150W-300W / LED Light source having atleast 4 LEDs with auxiliary lamp
4	It should have HD Image output such as DVI-D, RGBS Connectors, Y/C Connectors and Composite video connector as Video Out.	It should have HD Image output such as DVI-D, RGBS Connectors, Y/C Connectors and Composite video connector or any other HD video output as Video Out.
7	Weight of processor with light source should not exceed 13Kgs.	Point Deleted
11	It should have facility of extra illumination for more light apart from normal brightness control.	Point Deleted
B	High definition Video bronchoscope	High definition Video bronchoscope
3	Tip Deflection Up-Down (0): More/ Down 130 or more	Tip Deflection Up-Down (0): Up 180 or More/ Down 130 or more
4	Rigid Distal Width (mm): 5.8mm or less 4.2	Rigid Distal Width (mm): 6.1 mm or less
11	Scope PVE Connector should be rotatable by 180 Degree to prevent LG cable from coiling during procedure	Point Deleted
12	Atleast Four Remote switches should be available on Control Body for setting customized functions	Remote switches should be available on Control Body for setting atleast 4 customized functions.
D	Sony 21" or more color HD Medical grade Monitor	27" or more color HD Medical grade Monitor from Same OEM/Sony