



Date: - 03/03/2026

Corrigendum

Sub: Corrigendum in Notice Inviting Tender for “Supply of Echocardiographic Scanning Machine for Trauma & Emergency Department at AIIMS, Raipur

Ref: GTE NO: AIIMS/R/CS/TNE/410/GTE/2025, Dated: 03-10-2025.

In context to following corrigendum is issued against subject tender:

Technical Amendments

S. No.	Existing Specification	To be read us
Point 6:	Operating frequency of at least 1 to 15 MHz.	Request to amend to "1 to 22 MHz or broader/higher range".
Point 13:	Battery life (run time) shall be at least 2 (Two) hours ¹¹ .	Request to change to "1.5 hours or more".
3.	Addendum	System should have auto EF Calculation
4.	Addendum	System should have panoramic imaging
5.	Addendum	Request to add "Strain elastography and Contrast imaging".
6.	Addendum	System should have user define pre-sets

Note: All other terms & condition will be remain unchanged.

Stores Officer
AIIMS, Raipur (C.G.)