



अखिल भारतीय आयुर्विज्ञान संस्थान (छत्तीसगढ़)
(All India Institute of Medical Sciences, Raipur (Chhattisgarh)
Tatibandh, GE Road, Raipur-492 099 (CG)
www.aiimsraipur.edu.in

No. AIIMS/R/HS/2019-20/NEPH/216/18624

Date – 04/04/2020

Corrigendum

Sub: Corrigendum in Notice Inviting Tender for “Supply of Consumable item (Dialyzer) for Nephrology Department”, AIIMS Raipur.

Pre bid meeting Dated 20-03-2020

In context to following corrigendum is issued against subject:

S.no .	Page/Clause/ Point	Existing Entry	To be read as/Addenda
1	On pg. 18 Technical Specification Annexure-I	Dialyzer Low Flux with Polysulfone membrane with effective surface area of 1.3 Squaremeter, with inline Steam Sterilization.	Dialyzer Low Flux with Polysulfone membrane with effective surface area 1.2-1.4 square meter with inline steam sterilization method/Steam Sterilization method.

Note:

- All other terms & condition will remain unchanged.*
- Corrigendum must be uploaded along with technical bid as acceptance.*

V.Sitaramu
Stores Officer (H)
AIIMS, Raipur