



अखिल भारतीय आयुर्विज्ञान संस्थान, रायपुर (छ.ग.)  
**All India Institute of Medical Sciences, Raipur (Chhattisgarh)**  
Tatibandh, GE Road, Raipur-492 099 (CG)

No. AIIMS/R/CS/Pul\_Med/19-20/105/OT  
Tender ID- 2021\_IMSRP\_610499\_1

Raipur, Dtd:- 05/01/2021

Tender Name – Supply of Fractional Exhaled Nitric Oxide (FENO) Monitor for Dept. of Pulmonary Medicine at AIIMS Raipur.

**Corrigendum**

Group/ Point no of Technical Specification on Annexure - I	Existing Sentence	To be read as
Other features	It should meet all international safety standards US FDA and European CE for human applications.	It should meet all the international safety standard US FDA/ CE/BIS/ISI/ Make in India for human applications
Essential Accessories	Sensor life should last for three years or 3000 tests.	Sensor life should be 3 or more years and unlimited tests can be done during warranty period.

Note :-

1. The above changes have no financial implication.
2. Technical specification remains the same as earlier except for above corrigendum.
3. All other terms and condition will remain unchanged.

**Stores Officer**  
**AIIMS Raipur (CG)**