



अखिल भारतीय आयुर्विज्ञान संस्थान, रायपुर (छत्तीसगढ़)

All India Institute of Medical Sciences, Raipur (Chhattisgarh)

Tatibandh, GE Road, Raipur-492 099 (CG)

www.aiimsraipur.edu.in

Dated: 05/08/2026

Corrigendum

Tender Ref No. AIIMS/R/CS/CCU/CRRT/25/E-19184 (2025_IMSRP_891857_1), dated 31-12-2025

With reference to above CPPP Bid No., the following amendment is being issued: -

Clause/Page No.	Existing Specification	Amendment To be read as
Page No. 40 & Point No. 3.33	Should have an Optical sensor to differentiate between therapy modality	Machines should be able to differentiate between therapy modalities.
Page No. 40 & Point no. 3.35	Should have an RS232 Port for Data transfer and interface	Should have as RS232 Port or USB for Data transfer and interface
Page No. 39 & Point no. 3.2	Should be able to perform Pre and /or Post CVVH, CVVHD, pCVVHD, Pre or Post CVVHDF, TPE, HP, SCUF	Should be able to perform Pre and /or Post CVVH, CVVHD, pCVVHD, Pre or Post CVVHDF, TPE.

All other terms and condition will remain unchanged.

Stores Officer
AIIMS Raipur (C.G.)