

Technical Corrigendum

Tender ID No. AIIMS/R/CS/TNE/ECHO/GTE/2026, Dated: 30.06.2026,

With reference to above Tender ID. No, the amendments are as mentioned below:

Sl. No.	Existing Tender Clause	Corrigendum/Addendum
1	Operating frequency 1–19 MHz	Minimum operating frequency range covering at least 1–19 MHz or better
2	Dynamic range 165–180 dB and depth ≥ 30 cm	Dynamic range 165–180 dB or higher and depth ≥ 30 cm
3	The system must have dedicated calculation packages for cardiac, vascular and acute care applications including wall motion analysis and valve assessment.	Reframed as: “The system must have dedicated calculation packages for cardiac, vascular and acute care applications with wall motion analysis and valve assessment, including One Touch facility to calculate EF, VTI, IVC, B-line.”
4	15" monitor + ≥ 10 " touchscreen	The system should have a high-resolution medical-grade monitor of at least 15 inches with anti-glare coating and wide viewing angle. The system may have touchscreen functionality or an equivalent rapid-access control interface designed for point-of-care emergency imaging.
5	Drop-safe certification	System and probes shall be rugged and suitable for hospital use. Documentary evidence/certification, wherever available, may be submitted.
6	Transducer Point 5: Minimum three active transducer ports with electronic switching facility.	The system shall have a minimum of three active transducer ports with electronic switching integrated within the ultrasound system, without requiring a trolley-mounted connector.
7	Centreline marker	Centreline marker or equivalent needle guidance technology
8	Regulatory compliance BIS/IEC/CE/USFDA	System shall comply with applicable regulatory standards such as BIS and/or CE and/or US FDA along with applicable IEC electrical safety standards
9	Stabilizer/CVT	Suitable power protection (stabilizer/CVT/online UPS or manufacturer-recommended equivalent) may be provided wherever required.
10	Warranty	Two-year warranty and 8-year CMC as per tender clause.
11	TEE Probe Warranty	Two-year warranty and 8-year CMC as per tender clause.

Store Office (CS)
AIIMS Raipur (C.G.)