



अखिल भारतीय आर्युविज्ञान संस्थान , रायपुर छत्तीसगढ़
All India Institute of Medical Sciences, Raipur Chhattisgarh
Tatibandh, GE Road,
Raipur-492 099 (CG)
www.aiimsraipur.edu.in

No. AIIMS/R/CS/FMT/11/2021/GTE/

Date -09.06.2023

Corrigendum

**Sub: Corrigendum in Notice Inviting Tender for “Supply of Grossing Station” for the department of Forensic Medicine & Toxicology, AIIMS Raipur.
Tender Id: - 2023_IMSRP_753287_1 date 15.05.2023**

In context to following corrigendum is issued against subject:

S.N.	Page/Clause/Point	Existing	To be read as/Addenda
1.	Page no 38 , Section-VII Technical Specification Point -5	Should have electromechanical lifting system	Should have electromechanical lifting system, The table height should be able to fluctuate between 850 and 1200 mm and total height between 1950 and 2300 mm.

Note: -

*All other terms & condition will be remain unchanged.
Corrigendum must be uploaded along with technical bid as acceptance.*

**Sr.Procurement cum Store Officer
AIIMS Raipur**

