



अखिल भारतीय आयुर्विज्ञान संस्थान ,रायपुर) छत्तीसगढ़
(All India Institute of Medical Sciences, Raipur (Chhattisgarh)
Tatibandh, GE Road, Raipur-492 099 (CG)
www.aiimsraipur.edu.in

No. AIIMS/R/CS/Micro/HICC/18/098/OT

Date – 09/07/2018

Corrigendum

Sub: “ATP bioluminescence based hygiene monitoring system with refilling pack pens swabs”

In context to following corrigendum is issued against subject tender:

S.no.	Page/Clause/Point	Existing	To be read as
1.	Page-17 Clause–Technical specification Point – 4	Logarithmic 0-0.5 log 10 RLU	Logarithmic 0-5.0 log 10 RLU

Note: All other terms & condition will be remain unchanged.

(सुशील सोनबेर)
भंडार अधिकारी
अखिल भारतीय आयुर्विज्ञान संस्थान,
रायपुर (छ.ग.)