



अखिल भारतीय आयुर्विज्ञान संस्थान, रायपुर (छ.ग.)
All India Institute of Medical Sciences, Raipur (Chhattisgarh)
Tatibandh, GE Road, Raipur-492 099 (CG)

No. AIIMS/R/CS/Pul_Med/19/062/OT
Tender ID- 2019_IMSRP_494430_1

Raipur, Dtd:- 09/09/2019

Tender Name – Supply of Electric infusion Syringe Pump for Dept of Pulmonary Medicine at AIIMS Raipur.

Corrigendum

Group/ Point no of Technical Specification on Annexure - I	Existing Sentence	To be read as
Point - 02	Screen: 3.5" TFT color LCD, 16:9 format Brightness: 1-8 levels adjustable, default is level 4	Screen: 3"- 4" TFT/LCD, 16:9 format Brightness: 1-8 levels adjustable, default is level 4
Point - 07	Bolus rate should be programmable to 10 to 1000 ml/hr or more with infused volume display and one key press bolus, auto bolus as a advance feature with programmable bolous rate. Display of Drug directory of more than 200 drugs, customized and adjustable.	Bolus rate should be programmable to 50 to 1000 ml/hr or more with infused volume display and one key press bolus, auto bolus as a advance feature with manual/programmable bolus rate. Display of Drug directory of more than 200 drugs, customized and adjustable.

Note :-

1. Technical specification remains the same as earlier except for above corrigendum.
2. All other terms and condition will remain unchanged.

Stores Officer
AIIMS Raipur (CG)