



No.: AIIMS/R/CS/Pediatric Surgery/491/23/GTE/A/Corrigendum/

Dated: 09/12/2025

**Corrigendum**

Tender ID 2025\_IMSRP\_867021\_2

With reference to above GTE No., the following amendment is being issued:-

Page/Clause/Point	Existing Specification	To be read as
Page No. 39, Point No. 01	The angle of view ranging between 5-12 degree, The working channel is 5Fr for Instrument up to 4- 5 Fr	Pediatric Operating Cystourethroscope compact 8 Fr (distal Tip), 5-12 degree, one step dilation, working length length 13-15 cm with angled eye piece, fiber optic light transmission incorporated with 2 lateral irrigation ports and 1 working channel 5 Fr for instruments up to 4 Fr., Fiber optic light transmission incorporated.
Page No. 39, Point No. 02	High frequency cords (Monopolar) – 3 Qty.	High frequency cords (Monopolar) – 3 Qty. (same make as Cystoscope)
Page No. 39, Point No. 03	Ball electrode should be 5Fr and length should not less than 40 cm – 3 Qty.	Ball/Button electrode should be 5 Fr and length should not less than 40 cm – 3 Qty. (same make as Cystoscope)
Page No. 39, Point No. 04	Hook bugbee electrode, 3 Fr, 40 cm or more – 5 Qty.	Hook Bugbee electrode, 2.4-3 Fr, 40 cm or more – 5 Qty. (same make as Cystoscope)

All other terms and condition will remain unchanged.

**Stores Officer**  
**AIIMS, Raipur (C.G.)**