



Date: 11-08-2022

Corrigendum

Global Tender Ref. No.: AIIMS/R/CS/CRDO/20/15/GTE date: 18-07-2022.
Tender ID No.: 2022_IMSRP_701413_1
Tender Name : Global Tender supply & installation of supply & installation of Intravascular Ultrasound (IVUS) and Fractional Flow Reserve with IFR and Coregistration for Department of Cardiology at AIIMS Raipur

Sr. No.	NIT Page No./ Item No./ Clause No.	Existing sentence	To be read as
1.	Section- VII Technical Specifications Page No. 41 Point No. 1	The system should be the latest generation digital Intra-vascular ultrasound for 360° image evaluation of coronary lumen.	The system should be the latest generation Intra-vascular ultrasound for 360° image evaluation of coronary lumen.
2	Section- VII Technical Specifications Page No. 41 Point No. 2	Should be a Windows based system capable of accepting phased array transducer technology.	Should be a Windows based system capable of accepting phased array or mechanical transducer technology.
3	Section- VII Technical Specifications Page No. 41 Point No. 6	Should be accompanied with Flat Panel LCD display of >18" high quality monitor with keyboard, track ball and mouse.	Should be accompanied with Flat Panel LCD display of >18" high quality monitor with keyboard or touch screen, track ball or touch pad and mouse.
4	Section- VII Technical Specifications Page No. 41 Point No. 34	Should come with Automatic pull back device.	Should come with Automatic pull back device and manual pull back options.
5	Section- VII Technical Specifications Page No. 41 Point No. 38	Prices of all accessories like pullback device, IVUS catheter, guidewire etc, to be quoted separately for future purchase.	Prices of all accessories like pullback device, IVUS catheter (approx. 100 for 03 years), guidewire (approx. 100 for 03 years) etc, to be quoted separately for future purchase. Pull back device will not be considered for L1 decision.
6	T& C. Page No. 21 Point No.35	L1 firm will be decided on the basis of total price of item For (including all taxes & duty) Warranty 05 Year + CMC 05 year.	L1 firm will be decided on the basis of total price of item including all taxes & duty + 5 Years Warranty + 5 Years CMC + Price of all consumables (100 qty. IVUS Catheters and Guidewires) to be freezed for 03 years.
7	Page No. 2	Bid Submission End Date: 13-08-2022, 03.00PM	Bid Submission End Date: 22-08-2022, 03.00PM
8		Bid Opening Date: 17-08-2022, 03.30PM	Bid Opening Date: 24-08-2022, 03.30PM

*Note: 1. All other terms & Conditions will remain unchanged.
2. Corrigendum must be uploaded along with technical bid as acceptance.*

Sr. Proc. cum Stores Officer,
AIIMS, Raipur (C.G.)