

दिनांक:- 13.08.2022

AIIMS/R/CS/ENT/1158/22/GTE/corrigendum/

Corrigendum

Tender ID No.: 2022_IMSRP_692700_1

With Reference to above tender ID No., the following amendment is being issued for the tender "Supply of ENT Navigation System" for the Department of ENT at AIIMS Raipur:-

S.	Specification	Existing	To be read as
N.	Sr. No.		
1	1.j	The system must work on electromagnetic technology and it must have electromagnetic based dynamic referencing so that registration is not lost even if the patient moves and helps navigate malleable instruments.	The system must work on Optical and advanced wireless passive marker tracking technology or electromagnetic technology and it must have electromagnetic based dynamic referencing so that registration is not lost even if the patient moves and helps navigate malleable instruments.
2	2.b	The software should allow to create multiple surgeon profile and set procedures and settings as per the surgeon requirement.	Preferably The software should allow to create multiple surgeon profile and set procedures and settings as per the surgeon requirement.
3	3.a	The system should be provide with flat pannel electromagnetic generator which can be positioned under patient head and / or the system should be provide which can be integrated into head rest for OR tables	The system should be provide with flat pannel electromagnetic generator which can be positioned under patient head and / or side mount which can be fixed around the head or the system should be provide which can be integrated into head rest for OR tables
4	3.b	The system should be provide with the ENT instruments set of malleable instruments with an outer diameter down to 0.8 mm which are navigable under electromagnetic technology. The portable navigation provides a	The system should be provide with the ENT instruments set of malleable instruments with an outer diameter down to 0.8 mm/ 1.5 mm or 2mm which are navigable under electromagnetic technology. The portable navigation provides a

Annexure I