



No. AIIMS/R/CS/BB/BIOCH/23/2025/GTE

Date - 13/11/2025

Technical Corrigendum

Sub: Corrigendum in Notice Inviting GTE Tender for “Flow Cytometer for Department of Blood Bank & Biochemistry at, AIIMS Raipur.

Tender Id :-2025_IMSRP_872446_1 Dated: 07-08-2025

In context to following corrigendum is issued against subject:

S.no .	Page/Clause /Point	Existing Entry	To be read as/Addenda
1.	Page no.03	Qty. 01 Set	Qty. 02 Set.
2	Page no. 38 Technical Specification Point no. 08	Point No. 8: The system should be quoted with both sampler acquisition options of a single tube and at least 20 20-tube carousels for auto acquisition. Bidder may offer optional 48/96/384 well microtiter plates.	"The system should be quoted with both sampler acquisition options of a single tube and at least 20 20-tube carousels for auto acquisition. Bidder should also provide 48/96/384 well microtiter plates included."
3	Page no. 38 Technical Specification Point no. 12	Tough/LCD screen computer with at least 21" dual monitor with at least Core 14th Generation Intel® Core™ i7-14700F Processor, 16GB RAM or Higher, Bidder should provide at least one colour laser printer. Also, should be pre-installed with the latest compatible Windows OS and MS Office. A cost of ITB external hard disk is to be inclusive in the equipment cost	"The system should be provided with compatible workstation, one color laser printer. A cost of 1TB external hard disc is to be inclusive in the equipment cost."
4	Page no. 39 Technical Specification Point no. 23	Point No. 23: OEM should have its own Service Centre and Service Engineer based at Raipur/Chhattisgarh After- sales services should be available on 24 hours basis and we must be able to provide services within 24 hours.	OEM should have its own service centre in India and service engineer based at Raipur/Chhattisgarh or adjacent states of Chhattisgarh. After sales services should be available on 24 hours basis and they must be able to provide services within 24 hours either in-person or remotely."
5	Page no. 38 Technical Specification Point no. 7	Compensation setting before, during and after data collection	Automated and manual compensation setting before, during and after data collection because it will be more convenient for the users and for the multicolour immunophenotyping study.

6	Page no. 38 Technical Specification Point no. 11	The system must have LIS compatibility	"The system must have LIS or equivalent technology to be compatible with file transfer.
7	Page no. 38 Technical Specification Point no. 12	Tough/LCD screen computer with at least 21"dual monitor with at least Core 14th Generation Intel Core i7- 14700F Processor, 16GB RAM or higher, bidder should provide at least one colour laser printer. Also should be pre installed with latest compatible windows OS and MS office. A cost of 1TB external hard disc is to be inclusive in the equipment cost.	"The system should be provided with compatible workstation, one color laser printer. A cost of 1TB external hard disc is to be inclusive in the equipment cost."
8	Page no. 39 Technical Specification Point no. 21	Should be US FDA OR EUROPEAN CH-IVD or BIS approved product	Should be US FDA or European CE-IVD or BIS approved product". It will reduced the participation from others
9	Page no. 39 Technical Specification Point no. 23	OEM should have its own service centre and service engineer based at Raipur/Chhattisgarh. After sales services should be available on 24 hours basis and we must be able to provide services within 24 hours.	OEM should have its own service centre in India and service engineer based at Raipur/Chhattisgarh or adjacent states of Chhattisgarh. After sales services should be available on 24 hours basis and we must be able to provide services within 24 hours either in-person or remotely."
10	Page no. 40 General Technical Specification Point no. a	Five years Comprehensive Warranty as per Conditions of Contract of the TE document for complete equipment (including Batteries for UPS, other vacuumatic parts wherever applicable) Warranty period will be 5 years from the date of installation, commissioning and Site Modification Work from the date of satisfactory installation, commissioning, trial run & handing over of equipment to Hospital/Institution/Medical College.	Two years Comprehensive Warranty as per Conditions of Contract of the TE document for complete equipment (including Batteries for UPS, other vacuumatic parts wherever applicable) Warranty period will be 5 years from the date of installation, commissioning and Site Modification Work from the date of satisfactory installation, commissioning, trial run & handing over of equipment to Hospital/Institution/Medical College
11	Page no. 40 General Technical Specification Point no. 2 (a)	The cost of Comprehensive Maintenance Contract (CMC) which includes preventive maintenance including testing & calibration as per technical/ service /operational manual of the manufacturer, labour and spares, after satisfactory completion of Warranty period may be quoted for next 5 years on yearly basis for complete equipment (including Batteries for UPS, other vacuumatic parts wherever applicable) and Site Modification Work (if any). The supplier shall visit each consignee site as	The cost of Comprehensive Maintenance Contract (CMC) which includes preventive maintenance including testing & calibration as per technical/ service /operational manual of the manufacturer, labour and spares, after satisfactory completion of Warranty period may be quoted for next 8 years on yearly basis for complete equipment (including

		recommended in the manufacturer's technical/ service /operational manual, but at least twice in six months (i.e. 4 preventive maintenance/ year) apart from all breakdown visits, during the CMC period	Batteries for UPS, other vacuumatic parts wherever applicable) and Site Modification Work (if any). The supplier shall visit each consignee site as recommended in the manufacturer's technical/ service /operational manual, but at least twice in six months (i.e. 4 preventive maintenance/ year) apart from all breakdown visits, during the CMC period
--	--	---	---

Note:

- a) All other terms & condition will remain unchanged.*
- b) Corrigendum must be uploaded along with technical bid as acceptance.*

Store Officer (C)
AIIMS, Raipur