



अखिल भारतीय आयुर्विज्ञान संस्थान ,रायपुर (छत्तीसगढ़)
All India Institute of Medical Sciences, Raipur (Chhattisgarh)
Tatibandh, GE Road,
Raipur-492 099 (CG)
www.aiimsraipur.edu.in

No. AIIMS/R/CS/Micro/17/41/Tender/

Raipur, Dtd:- 14/05/2018

Corrigendum

Tender ID.No.2017_IMSRP_268825_2, Dated:02.04.2018.

With reference to above Tender ID.No., the following amendment is issued:

S.No.	Page no in the NIT	Existing Tender Specification	To be read as
1.	Page.No.19	Annexure 1 PCR Workstation: UV shelf with integrated pipette holder	UV shelf with integrated/free standing pipette holder.
		Conform of electrical safety standards IEC 60601-1,UL 61010-1,EN 61010-1.	Conform of electrical safety standards IEC 60601-1/UL 61010-1/EN 61010-1 as all are the same standard.
2.	Page.No.21	Annexure 2 Laminar Flow Material of construction :Inner – Stainless steel: Outer – Powder coated steel preferably with microbicidal coating.	Material construction : Inner – Stainless steel: Outer –Powder coated steel.
		Electrical Safety requirements UL 61010-1,EN 61010-1,IEC 61010-1.	Electrical Safety requirements UL 61010-1/EN 61010-1/IEC 61010-1.

All other terms and condition will remain unchanged.

(सुशील सोनबेर)
भंडार अधिकारी
अखिल भारतीय आयुर्विज्ञान संस्थान,
रायपुर (छ.ग.)