



अखिल भारतीय आयुर्विज्ञान संस्थान, रायपुर (छत्तीसगढ़)
All India Institute of Medical Sciences, Raipur (Chhattisgarh)
Tatibandh, GE Road,
Raipur-492 099 (CG)
www.aiimsraipur.edu.in

No. AIIMS/R/HS/MOT/06161/2020/RC

Dated: - 17-02-2021

Technical Corrigendum Notice

Tender ID. No. 2021_IMSRP_613562_1 **Dated:** 02/02/2021

With reference to above Tender ID. No., the following amendment is issued:

S. No.	Page No. Details	Existing Tender Specification	To be read as
1	Page No.13	<ul style="list-style-type: none">Cold, Bright White light –over 2500 LUX	<ul style="list-style-type: none">Cold, Bright White light –over 2500 LUX using UV diode and semiconductor technology.

In addition to above amendment, the following specifications of items that are needed to be revised which are as under:

S. No.	Page No. Details	Title under which amendments made	Revised additional specifications
1	Page No.13 [Annexure -1]	Technical Specification Of High End Fibre-optic LED Handle with Green System identification for Laryngoscopes (Adult) for Major Operation Theatre	<ul style="list-style-type: none">Atleast 3 year warranty on Blades for any manufacturing defects.Demonstration of the equipment may be asked for ensuring good quality.

All terms and conditions will remained same.

Stores Officer (H)
AIIMS Raipur (C.G)