



अखिल भारतीय आयुर्विज्ञान संस्थान ,रायपुर) छत्तीसगढ़  
(All India Institute of Medical Sciences, Raipur (Chhattisgarh)  
Tatibandh, GE Road, Raipur-492 099 (CG)  
[www.aiimsraipur.edu.in](http://www.aiimsraipur.edu.in)

No. AIIMS/R/CS/21149/Oph/19/OT

Date – 17/07/2019

### **Corrigendum**

**Sub: Corrigendum in Notice Inviting Tender for “Supply of Retina Laser(Multi Spot) with SL,IO and Endo Delivery System” for Department of Ophthalmology, AIIMS Raipur.**

**Tender Id: - 2019\_IMSRP\_480077\_1 date 21/06/2019**

In context to following corrigendum is issued against subject:

S.no.	Page/Clause/Point	To be added
1.	Page 19 (on Technical Specifications to be added)	<ul style="list-style-type: none"><li>• PAR FOCUS TECHNOLOGY</li><li>• <b>Insight view display:</b> Micro display integrated in LSL. Surgeon need not to cheque his parameters outside or by moving his eye during the laser procedure. Surgeon can stay focused &amp; can see all the parameters on the right eye piece.</li><li>• Fixed clear view PSF (physician safety Filter).</li><li>• <b>Automated laser treatment summary report:</b> Detailed reports are generated after successfully completion of patient's treatment. As a result surgeon receive a complete glance of Laser Treatment summary report in touch control panel.</li><li>• Smaller in volume lighter in weight fibre ports.</li><li>• Upgradation to selective Laser Trabeculoplasty (SLT) for Glaucoma</li></ul>

**Note:**

- All other terms & condition will be remain unchanged.*
- Corrigendum must be uploaded along with technical bid as acceptance.*

भंडार अधिकारी  
अखिल भारतीय आयुर्विज्ञान संस्थान,  
रायपुर (छ.ग.)