

# अखिल भारतीय आयुर्विज्ञान संस्थान ,रायपुर) छत्तीसगढ़ (All India Institute of Medical Sciences, Raipur (Chhattisgarh)

Tatibandh, GE Road, Raipur-492 099 (CG)

www.aiimsraipur.edu.in

#### No. AIIMS/R/CS/181/BPS/24/GTE

### Date - 16/10/2024

## Corrigendum

Sub: Corrigendum in Notice Inviting Tender for "Supply of ND Yag Laser with Accessories for the Department of Burn and Plastic Surgery at AIIMS Raipur.

Tender Id: - 2024\_IMSRP\_825236\_1 date 09/09/2024.

In context to following corrigendum is issued against subject:

Point.	Existing	To be read as/Amendment
No.	THE OVOTEM CHOILD HAVE	
3	THE SYSTEM SHOULD HAVE	THE SYSTEM SHOULD HAVE LONG PULSED
	LONG PULSED Nd: YAG, Q SWITCH ND: YAG, ERBIUM	ND: YAG, Q SWITCH ND: YAG, ERBIUM YAG & HAIR REMOVAL TECHNOLOGY WITH
	YAG & IN MOTION HAIR	INBUILT COOLING
	REMOVAL TECHNOLOGY	INBUILT COOLING
	WITH INBUILT COOLING	
4	THE MAXIMUM SPOT SIZE FOR	THE SPOT SIZE FOR Q SWITCH LASER
7	Q SWITCH SHOULD BE UPTO	SHOULD BE UPTO 10 MM
	10MM	SHOOLD BL OF TO IV WIN
5	IT SHOULD HAVE A FRACTIONAL 5MM X	IT SHOULD HAVE A FRACTIONAL WHICH IS
•	5MM TIP FOR Q SWITCH LASER	ATLEAST OF 5MM X 5MM TIP AND ABOVE
		FOR Q SWITCH LASER.
6	THE LONG PULSE ND: YAG HP	THE SPOT SIZE FOR MULTI SPOT LONG
	SHOULD COME WITH A 10MM	PULSED ND: YAG HAND PIECE <b>SHOULD BE</b>
	SPOT SIZE	UPTO 10MM.
7	IT SHOULD HAVE A	IT SHOULD HAVE A VASCULAR HAND
	VASCULAR HANDPIECE	PIECE RANGE OF <b>500-755NM</b> WITH
	RANGE OF 500 -600nm WITH	INBUILT COOLING OR THE LONG
	INBUILT COOLING	PULSE ND: YAG WHICH CAN DO
		VASCULAR TREATMENTS.
		VIIDEELIK IKLIIIVILIVIS.
8	IT SHOULD HAVE A MOTION	IT SHOULD HAVE <b>OPTION TO CHANGE</b>
	SENSON SCANNER IN ERBIUM	SPOT SIZE, SHAPE & SPOT DENSITY IN
	YAG.	FRACTIONAL LASER
9	ERBIUM SHOULD HAVE A PIXEL TIP	THE ERBIUM TIP SHOULD BE 2MM,
	7X7 MM	4MM, 9MM & ALSO A FRACTIONAL TIP
		FOR RESURFACING.
10	IT SHOULD ALSO HAVE	THE SURGICAL ERBIUM TIP SHOULD
10	SURGICAL ERBIUM TIP of 1mm and 4 mm	BE DIFFERENT SIZES UP TO 18MM (AT
	SURGICAL ERBIOWI THE OF THIM and 4 HIM	LEAST 3 SHOULD BE PROVIDED)
11	IT SHOULD HAVE ATLEAT 3	NO CHANGES
11		NO CHANGES
	INSTALLATION IN MAJOR	
	GOVERNMENT MEDICAL COLLEGES	
	ACROSS INDIA	
12	THE MANUFACTURER SHOULD HAVE	THE MANUFACTURER SHOULD HAVE A
	AN DIRECT OFFICE IN INDIA	DIRECT OFFICE IN INDIA OR FOR
		BETTER SERVICE THE COMPANY
		WHO IS DEALING WITH THE DEVICE
		SHOULD HAVE A DIRECT OFFICE
		WITH SERVICE ENGINEERS ACROSS
		INDIA

15	Accessories: Pixel Er: YAG 2940 nm Hand Piece Tip 1x7 Scanner Tip Pixel 7x7 1mmtip, 4mm tip	ALL REQUIRED HAND PIECES & ACCESSORIES SHOULD BE INCLUSIVE & NOT SEPARATE WHICH IS REQUIRED TO OPERATE THE DEVICE.
	Clear lift Q Switched Nd : YAG Laser Hand Piece 10 mm tip 4D Tip Fractionated tip (5*5) KTP Tip CVD Tip (1064) QS Scanner Tip leg Assy Plus 5 SHR 5cm	

#### Note:

- a) All other terms & condition will be remain unchanged.
- b) Corrigendum must be uploaded along with technical bid as acceptance.

Stores Officer AIIMS Raipur