



Date: - 18/02/2020

Corrigendum

विषय/Sub: Corrigendum in Notice Inviting Tender for “Supply of Mobile C-arm Image intensifier System” for Urology and Orthopaedics Department at AIIMS, Raipur

संदर्भ/Ref: Tender No. AIIMS/R/CS/URO AND ORTHO/18/40/OT/B

In context to following corrigendum is issued against subject tender:

Sr. No	Page/Clause/ Point	Existing Sentence	To be read as/ Addendum
01	Page 19, Point No. A. 1	2.Power output: 2 KW or more	Power output: 3 KW or more
02	Page 19, Point No. A. 5	5. mA in fluoroscopy: 0.1 to 4 mA or more in normal fluoroscopy and 12 mA or more in High Level Fluro	5. mA in fluoroscopy: 0.2 to 4 mA or more in normal fluoroscopy and 12 mA or more in High Level Fluro
03	Page 19, Point No. C. 2 (iv)	2. It should have the following range of movements: (iv) Swing / panning movement: +/- 12 degrees or more	2. It should have the following range of movements: (iv) Swing / panning movement: +/- 10 degrees or more
04	Page 20, Point No. H. 1	1. Equipment should have AERB Type Approval Certificate for radiation safety	1. The quoted model should be AERB type APPROVED and the vendor should get necessary approvals for functioning of the unit in the department soon after procurement
05	Page 20, Point No. K. 1	Nil	Added: Should provide warranty of 5 years and CMC of 5 years.
06	Page 20, Point No. L. 1	Nil	Added: All the firms are required to submit their catalogue along with tender/quotation clearly marking the quoted item and No. with a highlighter.
07	Page 20, Point No. M. 1	Nil	Added: Demonstration of the equipment may require for technical evaluation in case of discrepancy.

Note: All other terms & condition will be remain unchanged.

**Sr. Administrative Officer
AIIMS Raipur (CG)**