



No. AIIMS/R/CS/3602/OBGY/18/OT

Date – 19/02/2019

### Corrigendum

**Sub: Corrigendum in Notice Inviting Tender for “Supply of Gynae Operation Theatre Table” for Department of OBGY, AIIMS Raipur.**

**Tender Id: - 2019\_IMSRP\_436415\_1 date 29/01/2019**

In context to following corrigendum is issued against subject:

S.no.	Page/Clause/Point	Existing	To be read as/Addenda
1.	Page 19 <b>Title</b>	Multipurpose electro hydraulic with manual override mobile Table with divided leg section suitable for all Gynaecological surgical procedures, complete with 5cm mattress and corded handset.	Multipurpose electro hydraulic with manual override and <b>remote operated with electrical override</b> mobile Table with divided leg section suitable for all Gynaecological surgical procedures, Complete with 5cm mattress and corded handset.
2	Page 19 Point No. 2 General Operating Table Features	It should have 4 or 5 sections tabletop, which should be made of a special scratch resistant, hardwearing and easy to clean material. Base column cover to be made of 100% stainless steel alloy and stainless steel.	It should have <b>5 or 6</b> sections tabletop, which should be made of a special scratch resistant, hardwearing and easy to clean material. Base column cover to be made of 100% stainless steel alloy and stainless steel.
3	Page 19 Point No. 4 General Operating Table Features	It should have battery powered, with facility for connection to mains electricity for immediate use. Battery Exhaustion protection and low battery warning via an audible beep/display indicator should be available.	It should have battery powered, with facility for connection to mains electricity for immediate use. Battery Exhaustion protection and low battery warning via an audible beep/display indicator should be available <b>with battery backup upto 7 days or minimum 150 operations.</b>
4	Page 20 Technical Data	Cranial & caudal traversing: 200-300 mm. it should be operated with remote control	Cranial & caudal traversing: <b>150-150 mm and above.</b> it should be operated with remote control
5	Page 20 Technical Data	Maximum patient weight: 180 kg or more in all position.	Maximum patient weight: <b>225 kg</b> or more in all position.
6	Page 20 Point No. 2 Accessories Specification	Lithotomy leg holders “Allen type” (adult and paediatric)-1set each	Lithotomy leg holders “Allen type” ( <b>Gas Spring type</b> ) (adult and paediatric)-1set each

**Note:**

- All other terms & condition will be remain unchanged.
- Corrigendum must be uploaded along with technical bid as acceptance.

भंडार अधिकारी  
अखिल भारतीय आयुर्विज्ञान संस्थान,  
रायपुर (छ.ग.)