



Date: - 22/02/2021

Corrigendum

विषय/Sub: Corrigendum in Notice Inviting Tender for “Supply of “Refrigerated Blood Bag Centrifuge & Automated Cell Washer” for Transfusion Medicine & Blood Bank Department at AIIMS, Raipur

संदर्भ/Ref: Tender No. AIIMS/R/CS/BB/12/10/2020/OT

In context to following corrigendum is issued against subject tender:

Item No. 1. Refrigerated Blood Bag Centrifuge for Making Blood Components

Sr. No	Page/Clause/ Point	Existing Sentence	To be read as/ Addendum
01	Page 22, Point No.(1) a	1. Design a) Blood Bag centrifuge should be able to accommodate 16 fully filled double, triple, quadruple, Penta blood bags and blood bags with in-line filters for the purpose of component preparation.	1. Design a) Blood Bag centrifuge should be able to accommodate 12 or16 fully filled double, triple, quadruple, Penta blood bags and blood bags with inline filters for the purpose of component preparation.
02	Page 23, Point No.(17)	Nil	Addendum Electrical standards IEC 61010-2-020 3 rd Edition; IEC 61010-2-101 3 rd Edition, EN 61326-1 Class B, EN 61326-2-6, EN 62304, EN 62366, EN ISO 14971, EN ISO 13485

Item No. 2. Automated Cell Washer

Sr. No	Page/Clause/ Point	Existing Sentence	To be read as/ Addendum
01	Page 23, Point No.(3)	3. Maximum Capacity: 24 x 5 ml with tube size 10X75ml or 12X 75ml tube size.	3. Maximum Capacity: 24 tubes with tube size 10/X75mm or 12X 75mm
02	Page 23, Point No.(5)	5. DIMENSIONS (W x D x H): 330 x 480 x 280 mm.	5. DIMENSIONS (H x W x D): 33-43 x 35-48 x 25-45 mm

03	Page 23, Point No.(9)	9. Electrical Requirements 115V; 60Hz I 230V 50-60Hz	9. 220/230V 50/60Hz with Electrical standards IEC 61010-1, IEC 61010- 2-020, EN ISO 13485
04	Page 23, Point No.(13)	Nil	13. Addendum: Maximum admissible noise <55 dB

Note: All other terms & condition will be remain unchanged.

**Stores Officer
AIIMS Raipur (CG)**