



**Date: - 23/02/2019**

**Corrigendum**

**विषय/Sub:** Corrigendum in Notice Inviting Tender for “Supply of 30-W Holmium Laser Machine” for Urology Department at AIIMS, Raipur

**संदर्भ/Ref:** Tender No. AIIMS/R/CS/URO/18/26/OT

In context to following corrigendum is issued against subject tender:

Sr. No	Page/Clause/ Point	Existing Sentence	To be read as
01	Page 17, Point No. 1	Maximum power up to 30 watts (adjustable).	Maximum power output (adjustable) must be 30 watts or higher
02	Page 17, Point No. 2	Selection of at least energy settings between 0.3J-2.0J or more than 2.0J.	Selection of energy setting between 0.3J-3.5J or higher.
03	Page 17, Point No. 3	Selection of pulse frequency between 1 Hz to 15 Hz or more.	Selection of pulse frequency between 3 Hz of lower to 23 Hz or higher.
04	Page 17, Point No. 13	Nil	<b>Added:</b> Should provide evidence validating the delivery of optimum energy at fibre end (+/- 10%)
05	Page 17, Point No. 14	Nil	<b>Added:</b> Physical demonstration of the system must be arranged if required during technical evaluation.

Note: All other terms & condition will be remain unchanged.

**Stores Officer  
AIIMS Raipur (CG)**