



Date: - 23 /03/2021

Corrigendum

विषय/Sub: Corrigendum in Notice Inviting Tender for “Supply of Endoscopic Tissue Morcellator Set” for Urology Department at AIIMS, Raipur

संदर्भ/Ref: Tender No.AIIMS/R/CS/URO/20/31/OT

In context to following corrigendum is issued against subject tender:

Sr. No	Page/Clause/ Point	Existing Sentence	To be read as/ Addendum
01	Page 23, Point No. G. (1,2,3 & 4)	[G] Suction-Irrigation Pump Unit: Qty. 01 Set 1. It should be use as Suction device/Suction control device for Morcellationprocedure. Suction pump/Control unit for suction and continuous aspiration of irrigation fluid and enucleated tissue.	[G] Suction-Irrigation Pump Unit: Qty. 01 Set Controlled irrigation or suction unit for Urological Procedure to maintain the cavity pressure as desirable by Surgeon. Pressure of irrigation to be maintained between 0-150 mm hg. Flow 100-400 ml/min.
02	Page 23, Point No. G. (1,2,3 & 4)	[G] Suction-Irrigation Pump Unit: Qty. 01 Set 3. It should be controlled manually as well as by foot switch	[G] Suction-Irrigation Pump Unit: Qty. 01 Set Should consists of main unit touch screen panel, single roller pump, pre-configured procedure option based on software being installed, pump head, automatic tubing detection facility, controlled irrigation function.
03	Page 23, Point No. G. (1,2,3 & 4)	[G] Suction-Irrigation Pump Unit: Qty. 01 Set 4. It should be supplied with following accessories: (i) Power supply: 100-240 VAC, 50/60 Hz. (ii) Sterilizable Tubing Setfor suction: Qty. Two set (iii) A mainpower cable. (iv) Connecting cable, length 100cm. (v) It should have integrated tissue collection disposable kit: Qty. 20	[G] Suction-Irrigation Pump Unit: Qty. 01 Set It should be supplied with following accessories: (i) Power supply: 100-240 VAC, 50/60 Hz. (ii) Sterilizable Tubing Set: Qty. 1 set (iii) A mainpower cable. (iv) Disposable Tubing set – pack of 10

Note: All other terms & condition will be remain unchanged.

Stores Officer
AIIMS, Raipur (C.G.)