



Date: - 23/09/2019

Corrigendum

विषय/Sub: Corrigendum in Notice Inviting Tender for “Supply of Adult Lower tract instrument & Cysto-Resection (Monopolar and TURis) Set” for Urology Department at AIIMS, Raipur

संदर्भ/Ref: Tender No. AIIMS/R/CS/URO/18/13/B

In context to following corrigendum is issued against subject tender:

Sr. No	Page/Clause/ Point	Existing Sentence	To be read as/ Addendum
01	Page 17, Point No. 1	Forward oblique 12-degree telescope, 4 mm diameter, autoclavable, length 30 cm, fibre optic light transmission incorporated- 1pc	Forward oblique 12-degree telescope, 4 mm diameter, autoclavable, length 30 cm, fibre optic light transmission incorporated, with autoclave tray from same OEM-1 pc
	Page 17, Point No. 2	Forward oblique 30-degree telescope, 4 mm diameter, autoclavable, length 30 cm, fibre optic light transmission incorporated- 2pc	Forward oblique 30-degree telescope, 4 mm diameter, autoclavable, length 30 cm, fibre optic light transmission incorporated, with autoclave tray from same OEM-2 pc
	Page 17, Point No. 3	Oblique 70-degree telescope, 4 mm diameter, autoclavable, length 30cm, fibre optic light transmission incorporated- 1pc	Oblique 70-degree telescope, 4 mm diameter, autoclavable, length 30cm, fibre optic light transmission incorporated with autoclave tray from same OEM-1 pc
	Page 17, Point No. 8	Telescope bridge with single lockable working channel- 1pc	Telescope bridge with single and double lockable working channel- 1pc
02	Page 17, Point No. 10	Biopsy Forceps 5Fr and 7 Fr, double action jaws, flexible, length 30 cm or more- 1pc each	Compatible forceps 5fr and 7 fr., double action jaws, flexible, length, 30cm or more- 1pc each from any make reputed make.

Sr. No	Page/Clause/ Point	Existing Sentence	To be read as/ Addendum
	Page 17, Point No. 14	Toomey glass syringe- 100 ml & 50 ml - 1 pc each	Toomey glass syring-up to 150ml – 2pc
	Page 18, Point No. 15	Saline TUR Set: System should include the following: Bipolar HF-Cable- 1 pc	Bipolar HF-Cable 4pc
		Monopolar HF-Cable- 1 pc	Monopolar HF-Cable 4pc
03	Page 19, Point No. 15	Should be supplied with compatible electrodes: Bipolar Cutting Loop Electrode- 10 pc	Bipolar Cutting Loop Electrode- 12 pc
		Bipolar Collins/ Needle Electrode- 10 pc	Bipolar Collins/ Needle Electrode- 12 pc
		Bipolar Button Electrode for Plasma Vaporization- 10 pc	Bipolar Button Electrode for Plasma Vaporization- 12 pc
		Bipolar Enucleation Electrode- 10 pc	Bipolar Enucleation Electrode- 12 pc

Note: All other terms & condition will be remain unchanged.

वरिष्ठ प्रशासनिक अधिकारी
अखिल भारतीय आयुर्विज्ञान संस्थान
रायपुर (छ.ग.)