



क्रमांक: AIIMS/R/CS/Pedia S./18/256/OT/Corrigendum/

दिनांक: 23/10/2018

Tender ID No.: 2018_IMSRP_375924_1

With reference to above tender ID No., the following amendment is being issued:-

Corrigendum -05

Page/Clause/ Point	Existing	To be read as
Page no. 19, point no 16- i	Integrated EtCO2 monitoring on ventilator screen by main/side stream technique. It should be from the same make as ventilator. It should be on screen integrated, volumetric EtCO2 measurement with ability to display trend as well.	Integrated EtCO2 monitoring on ventilator screen by main/side stream technique.
New points	The equipment quoted should be latest and modern and should have been introduced within last 5 years.	Deleted
Add New Point		The machine should have facility for transpulmonary pressure monitoring system. Each machine should be provided with 2 esophageal pressure monitoring catheters with machine.

All other terms and condition will remain unchanged.

भंडार अधिकारी
अखिल भारतीय आयुर्विज्ञान संस्थान,
रायपुर (छ.ग.)