



अखिल भारतीय आयुर्विज्ञान संस्थान, रायपुर (छ.ग.)
All India Institute of Medical Sciences, Raipur (Chhattisgarh)
Tatibandh, GE Road, Raipur-492 099 (CG)

No. AIIMS/R/CS/Pul_Med/19/060/OT
Tender ID- 2019_IMSRP_509099_1

Dtd: 23/10/2019

Tender Name – Supply of USG and Echo Cardiology colour Doppler with Accessories for Department of Pulmonary Medicine at AIIMS Raipur.

Corrigendum

Subheading of Technical Specification on Annexure - I	Existing Sentence	To be read as
Transducers and Accessories Sl .No. 01	Should have 4 probe connectivity which should be microfit with flexible cables	Should have 3 or more probe connectivity which should be microfit with flexible cables.
Transducers and Accessories Sl .No. 02	6-13 MHz (± 1 MHz) multi-frequency, broadband linear array, 50 mm (+/- 5 mm) transducer for vascular, venous, small parts, musculoskeletal, Lung, nerve imaging with less than 40mm size. Higher frequency will be preferred	6-13 MHz (± 1 MHz) multi-frequency, broadband linear array, 50 mm (+/- 5 mm) transducer for vascular, venous, small parts, musculoskeletal, Lung, nerve imaging with less than 40mm size. Higher frequency will be preferred. Additional probes for cardiac, and convex also to be supplied.
Transducers and Accessories Sl .No. 07	UPS Compatible with USG Machine for minimum back up to 01 Hour	Stabilizer compatible with USG to be SUPPLIED to control Electrical Fluctuation.

Note :-

1. Technical specification remains the same as earlier except for above corrigendum.
2. All other terms and condition will remain unchanged.

Sr. Administrative Officer
AIIMS Raipur (CG)