



क्रमांक AIIMS/R/CS/Pul-Med/17/4977/OT/A

रायपुर दिनांक :-23/12/2017

**Corrigendum**

**Tender ID.No.2017\_IMSRP\_275969\_1, Dated: 04.12.2017.**

With reference to above Tender ID.No., the following amendment is issued:

Page/clause /point	Existing sentence	To be read as
NIT PAGE 19/3.4.t	Titration device should be as follows:- Modes: CPAP, Auto CPAP, S, Auto Bi-level, S/T, T, PC, AutoSV, AVAPS-AE. Weight Approximately: less than 2.5 kg Altitude compensation: Automatic Pressure ranges CPAP : 4- 20 cm H2O Auto CPAP : 4 -20 cm H2O S Mode :EPAP: 4-25 cm H2O, IPAP: 4 – 30 cm H2O Auto Bi-level : EPAP: 4 -25 cm H2O,IPAP 4-25 cm H2O S/T, T, PC:EPAP:4- 25 cm H2O, IPAP:4 – 30 cm H2O AutoSV :EPAP: 4-30 cm H2O, Pressure Support: 0- 26 cm H2O AVAPS-AE: 4-30 cm H2O, Pressure Support 2 – 26 cm H2O It should have integrated Humidifier It should be USFDA approved.	Titration device should be as follows:- Modes: CPAP, Auto CPAP, S, Auto Bi-level, S/T, T, PC, AutoSV, Hybrid ventilation with auto EPAP and auto breath rate. Weight Approximately: less than 2.5 kg Altitude compensation: Automatic Pressure ranges CPAP : 4- 20 cm H2O Auto CPAP : 4 -20 cm H2O S Mode :EPAP: 4-25 cm H2O, IPAP: 4 – 30 cm H2O Auto Bi-level : EPAP: 4 -25 cm H2O,IPAP 4-25 cm H2O S/T, T, PC:EPAP:4- 25 cm H2O, IPAP:4 – 30 cm H2O AutoSV :EPAP: 4-30 cm H2O, Pressure Support: 0- 26 cm H2O AVAPS-AE: 4-30 cm H2O, Pressure Support 2 – 26 cm H2O It should have integrated Humidifier It should be USFDA approved.
NIT PAGE 20/10	<b>10: Additional Accessories:</b> <ul style="list-style-type: none"><li>• Actigraphy-1</li><li>• LoFlo ( Etco2) Starter Kit- 1</li><li>• Standard set of sensor kit-2</li></ul>	<b>10: Additional Accessories:</b> <ul style="list-style-type: none"><li>• Actigraphy-1</li><li>• Stand alone ETCO2 monitor with vital sign monitoring facility</li><li>• Standard set of sensor kit-2</li></ul>

All other terms and condition will remain unchanged.

Vetted by TSC Member.

(सुशील सोनबेर)  
भंडार अधिकारी  
अ.भा.आ.सं.- रायपुर (छ.ग.)