



# अखिलभारतीयआयुर्विज्ञानसंस्थान, रायपुर (छत्तीसगढ़)

ALL INDIA INSTITUTE OF MEDICAL SCIENCES,  
TATIBANDH, GE ROAD,  
RAIPUR (C.G.) : 492099  
[www.aiimsraipur.edu.in](http://www.aiimsraipur.edu.in)

No. AIIMS/R/CS/OBGY/17/2565/Tender

Raipur, Dtd:- 24.07.2018

## Amendment in the issued Corrigendum

**Tender ID.No.2018\_IMSRP\_344767\_1, Dated:04.06.2018.**

With reference to above Tender ID.No., the following amendment is to be issued on published corrigendum

| S/n o. | Point No. | Existing Tender Specification   | To be read as  |
|--------|-----------|---|--|
| 1      | 5         | It should have endo mode for all endoscopic work.   | 1.It should have endo mode for all endoscopic work , <b>Bipolar cut for use with other bipolar laparoscopic instruments,</b><br><b>2.It must have option of Bipolar resection software upgradation and must be compatible with resectoscopes from different manufacturers.</b>   |
| 2      | 12        | It should have reusable vessel sealing forceps required to seal upto 6-7mm vessel for open surgery and laparoscopy surgery (4 each)   | 1.It should have reusable vessel sealing instrument <b>with cutting</b> required to seal & <b>cut upto 7mm vessel</b> for open surgery and laparoscopy surgery ( <b>3 each</b> ).<br><b>2.7mm vessel Sealing should be approved by BIS / ISO/EUROPEAN CE / USFDA.</b><br><b>Relevant document has to be enclosed.</b>  |
| 3      | 13        | Should have bipolar scissor for open surgery and laparoscopy surgery (4 each)   | deleted  |
| 4      | 15        | <b>The system should be supplied with all the standard reusable accessories such as- Electrode set of 10 Nos. each (Pin point, flat tip-short, flat tip- long blade), Monopolar double pedal foot switch, earth pad – 5, bipolar forcep with cord -5, electrode handle with finger switch- 10 and other standard accessories</b><br><b>It should be comfortable for under water surgery like hysteroscopy</b><br><b>Open surgery reusable curved vessel sealing clamp-2</b><br><b>Laparoscopic reusable 5mm vessel sealing instrument-1</b> | The system should be supplied with all the standard reusable accessories such as- Electrode set of 10 Nos. each (Pin point, flat tip-short, flat tip- long blade), Monopolar double pedal foot switch, <b>Bipolar Single Pedal Foot switch earth pad – 50</b> , bipolar forcep with cord of following dimensions:<br><b>Straight Forceps (3 units):</b><br>• Blunt, straight tip of 1mm having forcep length of 145mm<br>• Fine, Short angled tip of 0.7mm diameter having forcep length of 165 mm<br>• Blunt, Straight tip of 2mm diameter having forcep length of 190 mm<br><b>Angled Forceps (1 unit):</b><br>• Fine, straight tip of 0.7mm diameter having length of 6mm & forcep length of 105mm and angulation of 170*<br><b>Bayonet Forcep (1 unit):</b><br>• Blunt, straight tip of 1.2mm diameter having length of 6mm & forcep length of 230mm<br>electrode handle with finger switch- 10 and It should be comfortable for under water surgeries like hysteroscopy |

|   |    |   |  |
|---|----|---|--|
| 5 | 16 | The unit should be having quality standard Certified by DIN ISO 9001/ EN 46001.                               | 1.The unit should be having quality standard certified by DIN ISO 9001/ EN 46001.<br><br>2.Theequipment must also comply mandatory with the requirement of the medical devicedirective of class I equipment and electromagnetic compatibility Certificatessuch as – IEC 6060 1-2 ( Certificate for standard Electro Magnetic Compatibility) IEC 6060 1-1 (Certificate for standard safety requirement for medical electrical Equipment) IP-X-8 ( 100% washable foot switch). |
| 6 | 17 | The equipment should be BIS/European CE/ US FDA approved and all accessories should be from same manufacturer | The equipment should be <b>BIS/ISO/European CE/USFDA</b> approved and all accessories should be from same manufacturer   |
| 7 | 20 | It should be provide trolley.   | Deleted  |

All other terms and condition will remain unchanged.

Vetted By:

Dr. Sarita Agrawal  
Professor  
Department of Obs & Gynae  
AIIMS Raipur

Dr. Debajyoti Mohanty  
Additional Professor  
Department of Surgery  
AIIMS Raipur

Dr. Nilajkumar Bagde  
Associate Professor  
Department of Obs.and Gynae.  
AIIMS Raipur

(सुशील सोनबेर)  
भंडार अधिकारी  
अखिल भारतीय आयुर्विज्ञान संस्थान,  
रायपुर (छ.ग.)