

Technical Corrigendum

Tender ID No. AIIMS/R/CS/PM/GTE/2025, Dated: 25.08.2025,

With reference to above Tender ID. No, the amendments are as mentioned below:

Name of the item	Contents of the representation	Whether accepted or not accepted	Reason for acceptance/non-acceptance	Whether corrigendum is required to be issued
CYRO Workstation	Footswitch with facility for swapping between programs - 2 Nos.	Accepted	This item is already included in the original technical specification under essential accessories	No
	Disposable Patient plate with equipotential ring - 250 Nos.	Not-accepted	The requirement of 100, as per the original tender specification, is sufficient	No
	Filter integrated Argon Plasma Coagulation flexible probe (side fire) - 20 Nos.	Accepted	Original tender specification mentioned ten. An additional ten will be useful for uninterrupted patient care	Yes
	Filter integrated Argon Plasma Coagulation flexible probe (Straight fire) - Delete this point	Not-accepted	This is item is essential	No
	Filter integrated Argon Plasma Coagulation flexible probe (circumferential fire) - Delete this point	Not-accepted	This is item is essential	No
	Monopolar cable for Endoscopic instruments - 10 Nos	Accepted	Additional monopolar cables will ensure uninterrupted patient care	Yes
	Argon Gas Cylinder for APC Unit with 3-5 Liters Capacity - 02	Accepted	This item is already included in the original technical specification under essential accessories	No
	Cryo Single use probes- 1.1mm-30 Pcs, 1.7mm -25 Pcs, 2.4mm - 25 Pcs	Accepted	Additional cryo probes will support uninterrupted patient care for longer durations	Yes