



**Date: - 24/12/2018**

### **Corrigendum**

**विषय/Sub:** Corrigendum in Notice Inviting Tender for “Supply of 2D Endovision System for Endourology & Laparoscopy” for Urology Department at AIIMS, Raipur

**संदर्भ/Ref:** Tender No. AIIMS/R/CS/URO/18/17/OT

Tender ID No.: 2018\_IMSRP\_408603\_1

In context to following corrigendum is issued against subject tender:

<b>Sr. No</b>	<b>Page/Clause/ Point</b>	<b>Existing Sentence</b>	<b>To be read as</b>
01	Page 17, Point no. 4 (b)	Narrow Band Imaging (NBI)-special filter light for observation of capillary vessels and fine patterns in the superficial layer of mucosa for early detection of lesions.	Camera should be have special filtered light for observation of capillary vessels and fine patterns in the superficial layer of mucosa for early detection of lesions.
02	Page 17, Point no. 9	Video Outputs: DVI/ HD-SDI/ Equivalent HD.	Video Outputs: DVI/ Equivalent Full-HD outputs
03	Page 17, Point no. 12	Camera module should be compatible for future use with Flexible Video Ureterorenoscope	Compatible Single Camera Module/ separate Camera Modules should be provided for use of both Flexible Fibre Ureterorenoscope and Flexible Video Ureterorenoscope
04	Page 18, Point no. (C) 1	LED based cold light source.	LED/XENON 300 watts light source (one XENON spare lamp of 300 watts should be provided
05	Page 18, Point no. (C) 6	Facility of switching between white light and Near Infra-Red light (IR) for ICG HD and special filter light (NBI) for observation of capillary vessels and fine patterns in the superficial layer of mucosa for early detection of lesions.	Facility of switching between white light and Near Infra-Red light (IR) for ICG HD and special filtered light for observation of capillary vessels and fine patterns in the superficial layer of mucosa for early detection of lesions
06	Page 19, Point no. (D) 8	Tubing (1 meter or longer) and connectors for gas cylinder and central gas supply with wrenches/ devices essential to make the connections.	Attachment hoses (1 meter or longer) and pin indexed connectors for gas cylinder and central gas supply with wrenches/ devices essential to make the connection to insufflator

<b>Sr. No</b>	<b>Page/Clause/ Point</b>	<b>Existing Sentence</b>	<b>To be read as</b>
07	Page 19, Point no. (F) 4	The quoted 26-inch or more panel flat monitor must be mounted on a mobile monitor arm for adjustment of position of the monitor with 2 articulations on the arm.	The quoted 26-inch or more flat panel monitor must be mounted on a mobile monitor arm with facility of adjustment of position of the monitor.

Note: All other terms & condition will be remain unchanged.

**Stores Officer  
AIIMS Raipur (CG)**