



AIIMS/R/CS/Neuro/18/252/OT /corrigendum

दिनांक:- 25.06.2019

Corrigendum

Tender ID No.: 2018_IMSRP_471486_1

With Reference to above tender ID No., the following amendment is being issued for the tender “Neuro ICU Monitor and Central Nursing Monitoring system” for Neurosurgery Department at AIIMS Raipur:-

Annexure I

Existing			To be read as
Schedule A			
S. no.	Specification Sr. No.		
1	2.	Monitor should have 15” TFT/LCD display with full touch screen user interface.	Monitor should have 15” or more TFT/LCD display with full touch screen user interface.
2	3.	Should be capable of displaying 10 or more waveforms.	Should be capable of displaying 8 or more waveforms.
3	17.	Monitor must be upgradable to Hemodynamic graphs to view Hemodynamic conditions of patients based on Frank-Starling principle and EGDT protocol.	Deleted
4	28.	Monitor should be upgradable to provide remote viewing of real time waveforms through internet.	Monitor should be upgradable to work with Central Nursing Station (CNS) and then CNS should provide remote viewing of real time waveforms through internet.
5	29.	The equipment should be BIS/ European CE/ USFDA Approved.	The equipment should be BIS/ (European CE & USFDA) Approved.
Schedule B			
1	12	The system should have facility to view the data remotely on I-pad or I-phone. It should be supplied with at least one I-pad	The system should have facility to view the data remotely on I-pad/I-phone/ laptop. It should be supplied with I-pad or Laptop.

भंडार अधिकारी
अखिल भारतीय आयुर्विज्ञान संस्थान,
रायपुर (छ.ग.)