



Date: 26-08-2017

CORRIGENDUM

Tender No. AIIMS/R/CS/Hospital/17/707/OT
Tender for “Supply & Installation of Basic Fowler Bed”
for Hospital at AIIMS-Raipur.

S. NO.	Page No./ Clause No./ Para No.	Existing sentence	To be read as/Addenda
1	Page No.14 Annexure-I	All MS parts should be pretreated and Epoxy powder coated with 60 micron thickness.	All MS parts should be pretreated with 9 tank process and Epoxy powder coated with not less than 60 micron thickness.
2		Should have provision of IV pole.	Should have provision of IV pole at four corners.
3		Should have removable polymer moulded Handle.	Should have removable stainless steel/ polymer moulded Handle.
4	Page No. 2, Clause No. 11	Tenderer must provide evidence of having supplied government hospital/ reputed private hospital organizations in India similar nature of items of at least ₹86,00,000.00 for Supply and Fitting (Installation) of Museum Rack and Cupboard of tender value in the last three years and the copy of the same should be uploaded.	Tenderer must provide evidence of having supplied government hospital/ reputed private hospital organizations in India similar nature of items of at least ₹86,00,000.00 for Supply and Installation of Basic Fowler Bed of tender value in the last three years and the copy of the same should be uploaded.
5	Page No. 1	Bid Document Download / Sale End Date: 29-08-2017 at 3.00 PM	Bid Document Download / Sale End Date: 05-09-2017 at 3.00 PM
6		Bid Submission End Date: 29-08-2017 at 3.00 PM	Bid Submission End Date: 05-09-2017 at 3.00 PM
7		Bid Opening Date : 30-08-2017 at 3.30 PM	Bid Opening Date : 06-09-2017 at 3.30 PM

Note:

1. All other terms & Conditions will remain unchanged.
2. Corrigendum must be uploaded along with technical bid as acceptance.

Stores Officer,
Central Store,
AIIMS Raipur